

# Equipment Planning

## Cherokee Nation Health Center WPM STILWELL EXPANSION PROJECT

Prepared for Childers Architecture  
December 6, 2019

Prepared by:  
Interior Logistics



**CPCI** = CONTRACTOR PROVIDED-CONTRACTOR INSTALLED

**OPCI** = OWNER PROCURED-CONTRACTOR INSTALLED

**OPVI** = OWNER PROCURED-VENDOR INSTALLED

**OPOI** = OWNER PROCURED-OWNER INSTALLED

**OPVICI** = OWNER PROCURED-VENDOR & CONTRACTOR INSTALLED

**OPSVI** = OWNER PROCURED-SPECIALTY VENDOR INSTALLED

**OPSVCI** = OWNER PROCURED-SPECIALTY VENDOR & CONTRACTOR INSTALLED



# Equipment Planning-Stilwell Expansion Project

**DEPTNO**            **00**  
**DEPTTITLE**        **PUBLIC FACILITIES**  
**ARCH RMNO**        **01-00-00**  
**RMTITLE**            **VESTIBULE, MAIN ENTRY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	9996	2	Bench, Sitting Suitable for Location	floor	Approx 48"w x 20"d x 20"h	None

**ARCH RMNO**        **01-00-01**  
**RMTITLE**            **LOBBY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1024	2	Seating, 3 Person, Couch Type	floor	84x30x28	None
OPVI	1026	2	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1052	1	Clock, Decorative, Battery Operated, Size TBD By Design	wall	24diam x3	Battery Op
OPVI	1303	2	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	Varies per Space	None: Consider Backing
OPVI	1410a	4	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3044	2	Wastebasket, Decorative Type, Fiberglass, Approximately 15 gal	floor	16"diamx38h	None
OPCI	3909	1	Kiosk Center, Check In: Informational (???FUTURE PURCHASE)	floor	24"w x 42"h x 24"d	E 115V; Data Com
OPOI	8202	1	Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x12"	None

**ARCH RMNO**        **01-00-02**  
**RMTITLE**            **REGISTRATION/RECEPTION 1ST FLOOR**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	3	Chair, Rotary, Std.Back,5- Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPOI	2941	3	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None

OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	3760	1	Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3903	3	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3936b	3	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	3	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-00-03**

**RMTITLE RESTROOM, WOMENS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO 01-00-04**

**RMTITLE RESTROOM, MENS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO 01-00-06**

**RMTITLE WAITING, PUBLIC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	5	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	29	Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	24x24x32	None
OPVI	1024	4	Seating, 3 Person, Couch Type	floor	84x30x28	None
OPVI	1026	16	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1410a	11	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	1421	2	Table, Sofa Suitable for Lounge/Waiting Areas	floor	36" diameter x 24"h	None

OPVI	3043	4	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4136	2	Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

**ARCH RMNO 01-00-07**

**RMTITLE LACTATION ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1005	1	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1754	1	Casework, Modular, Designed to Meet Space Req: LACTATION ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	2294	1	Diaper Pail, with Step	floor	12"w x 10"d x 32H	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
CPCI	3135	1	Diaper Changing Station, Horizontal Drop Down	wall	22"d x35.5"l x5.25"h	None

**ARCH RMNO 01-00-09**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-00-11**

**RMTITLE CIRCULATION, CORRIDOR (BY PHARMACY)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	8202	1	Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x12"	None



**DEPTNO** 01  
**DEPTTITLE** MEDICAL RECORDS  
**ARCH RMNO** 01-01-00  
**RMTITLE** RECEPTION (2 STAFF)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0152	2	Chair, Rotary, Bariatric, Std.Back, On Casters, Adjustable Arms	floor	TBD	None
OPVI	0430d	2	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPOI	2941	2	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	3760	1	Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3907	2	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936a	2	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	2	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO** 01-01-01  
**RMTITLE** OFFICE, FUTURE CHS

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0431d	1	Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x18"d x29"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm



OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1 Robe Hook, Double	wall	6x6x3	
OPOI	3907	1 Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936b	1 Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

**ARCH RMNO**            **01-01-02**  
**RMTITLE**             **OFFICE, CHS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0431d	1	Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x18"d x29"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3907x	1	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936b	1	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

**ARCH RMNO**            **01-01-03**  
**RMTITLE**             **OFFICE, CHS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None



OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm

**ARCH RMNO 01-01-06**

**RMTITLE OFFICE, OPEN (X5 CLERICAL STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	5	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	5	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVICI	1500I	1	Workstation,Open Office Furniture System,Modular, 5 staff	floor	See Design	E Quad x 8; data/com x4
OPOI	2941	5	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3907	5	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936a	5	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	5	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-01-07**

**RMTITLE CODER (X3 STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	3	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVICI	1500d	1	Workstation,Open Office Furniture System,Modular, 3 person	floor	Various	E Quad x 3: Data Com x 3
OPOI	2941x	3	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3907x	3	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com

OPOI	3936a	3	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	3	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-01-08**

**RMTITLE OFFICE, MEDICAL RECORDS STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0434d	1	Cabinet, Filing,Lateral, 4 drawer	floor	36x18x54	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-01-09**

**RMTITLE OFFICE, MEDICAL RECORDS STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0434d	1	Cabinet, Filing,Lateral, 4 drawer	floor	36x18x54	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO**            **01-01-10**  
**RMTITLE**             **STORAGE, FILE**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0435d	4	Cabinet, Filing, Lateral, 5 high- (optional flip top door for 5th space), Locking	floor	36"w x18.25"d x65.75"	None
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None
OPVI	1413d	1	Table, Work, Rectangular, Static, Standing Height	floor	48"l x18"d x36"h	None
OPVI	3039b	0	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**DEPTNO** 02  
**DEPTTITLE** EMPLOYEE FACILITIES  
**ARCH RMNO** 01-02-01  
**RMTITLE** LOUNGE, STAFF (EMPLOYEE AREA)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1005	8	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1411	2	Table, Square; SIZE & QTY TBD	floor	36x36x30	None
OPVI	3062	2	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp; Water Supply; NEMA 5-5P
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	33"w-doors open 36.5" x 33 3/8"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	2	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.75"h x24"w x 19.43"d	E 120V: 13.9 amp

**ARCH RMNO** 01-02-02  
**RMTITLE** RESTROOM, MENS

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO** 01-02-03  
**RMTITLE** RESTROOM, WOMENS

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO** 01-02-08  
**RMTITLE** LOUNGE, STAFF

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1005	12	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None

OPVI	1411	3	Table, Square; SIZE & QTY TBD	floor	36x36x30	None
OPVI	3062	2	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp; Water Supply; NEMA 5-5P
OPVI	3269d	2	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	33"w-doors open 36.5" x 33 3/8"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	2	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.75"h x24"w x 19.43"d	E 120V: 13.9 amp

**ARCH RMNO 02-02-07**

**RMTITLE WELLNESS CENTER, STAFF (FUTURE PURCHASE)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3083	1	Station, Sanitizing Suitable for Gym Areas; Wipe & Hand Sanitizer Dispenser, Waste Container	floor	47" H x 12" W x 17" D	None
OPVI	7103	1	Bike, Exercise, Upright, With Monitor: Precor (unit 169 lbs)	floor	42lx24wx54h	E Ded 120V; 15 amp; Data
OPVI	7103c	1	Bike, Exercise, Recumbent; No step thru With Monitor: Precor (unit 219 lbs)	floor	63"l x23"w x54"h	E Ded 120V; 15 amp; Data
CPCI	7110c	1	Mirror(s), Wall (Size, Design Determined by Architect)	wall	TBD: SEE ARCHITECTURAL	None
OPVI	7128b	2	Treadmill, Individual Walking Exercise, W/Handrails	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data-Com
OPVI	7129a	1	Exercise Unit, Elliptical Trainer, w/progress monitor	floor	84"lx 31"w x67"	E 120V Data-Com
OPVI	9996a	2	Bench, With 3 Cubbies, Suitable for Shoe Changing Areas in Gym	floor	36" Long x 13-1/2" Deep x 20" High	None

**DEPTNO**            **03**  
**DEPTTITLE**        **FACILITIES MAINTENANCE**  
**ARCH RMNO**        **01-03-00**  
**RMTITLE**            **OUTDOOR EQUIPMENT STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO**        **01-03-01**  
**RMTITLE**            **FACILITIES MAINTENANCE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0212	80	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	TBD	None
OPVI	0213	15	Dolly, Chair, Portable	floor	TBD	None
OPVI	2201	1	Cabinet, Storage: Flammable Liquid: up to 30 gal: self closing door	floor	43"w x 18"d x 44"h	None
OPVI	2256b	20	Shelving Unit, Open Wire Shelves, Static	floor	TBD	None
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x36"h	None
OPVI	3047	1	Trash Collector, 1/2 Cubic Yd, poly construction,color TBD on wheels with lid, color to be selected	floor	53lx27wx34h	None
OPVI	3047a	2	Trash Collector w/Lid,and Dolly, 44 gallons (color TBD by Function)	floor	21.5" diameter x36"	None



**DEPTNO** 04  
**DEPTTITLE** ENVIRONMENTAL SERVICES: HOUSEKEEPING

**ARCH RMNO** 01-04-01  
**RMTITLE** BIO-HAZARD

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3032	1	Container, Bio Hazard; Rated-Med Waste: Hinged Lid; Red 65 Gal Cap; mobile	floor	32.3"l x 25.3"w x 41.8"h	None

**ARCH RMNO** 01-04-02  
**RMTITLE** EVS CLOSET

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	2256b	1	Shelving Unit, Open Wire Shelves, Static	floor	TBD	None
OPVI	3046	1	Trash Collector, 32 Gal Brute, On dolly, Round with wrap around caddy for janitorial services	floor	TBD	None
OPVI	3047a	1	Trash Collector w/Lid,and Dolly, 44 gallons (color TBD by Function)	floor	21.5" diameter x36"	None
OPVI	3079	1	Bucket, Mop with ringer	floor	20"w x 28"l x 20"h	None

**ARCH RMNO** 01-04-03  
**RMTITLE** EVS BULK STORAGE

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	2254	4	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	2256b	1	Shelving Unit, Open Wire Shelves, Static	floor	TBD	None
OPVI	3046	2	Trash Collector, 32 Gal Brute, On dolly, Round with wrap around caddy for janitorial services	floor	TBD	None

**ARCH RMNO** 01-04-04  
**RMTITLE** OFFICE, EVS

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None

OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3913	1	Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm

**ARCH RMNO 01-04-05**

**RMTITLE LINEN, CLEAN**

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI	3005	2	Cart, Linen, w/ 4 shelves and Staph Check enclosure	floor	36"w x 21"d x 64"h	None
------	------	---	---	-------	--------------------	------

**ARCH RMNO 01-04-06**

**RMTITLE DRYER, LAUNDRY**

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI	1413e	1	Table, Work, Rectangular, Static, Standing Height	floor	36"l x24"d x36"h	None
OPCI	3021	1	Dryer, Front Load, Commercial, Electric, 7 Cu Ft Capacity	floor	26.88"w x28" install depth x44.34"h	E 240/60/1, 3-prong cord, 30 AMP; SEE ADD DETAILS
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-04-07**

**RMTITLE WASHER, LAUNDRY**

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI	3017	1	Hamper, Laundry: round, mobile	floor	24"diameter x 34"h	None
OPCI	3020	1	Washer, Extractor, Front Load, Commercial, Energy Star Rated, 20 lb capacity	floor	27"w x27.6"d x42.5"h	E 208-240/60/1, 2W+G,20AMP; SEE ADDITIONAL DETAILS

OPVI	3023	1 Dispensing System, Laundry Detergent (LIKELY TO BE SOURCED THRU OWNER	wall mount	19.75"l x 6.75"w x29"h	E 115V?; Water Supply
OPVI	3062	1 Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer	wall	TBD	None

**DEPTNO** 05  
**DEPTTITLE** MEP SPACE  
**ARCH RMNO** 01-05-01  
**RMTITLE** MEP LARGE SPACE, 1ST FLOOR

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI 3045 1 Wastebasket, 25 gallon, with rigid Liner, UL rated, floor 18" diameter x 42"h None  
 OPVI 3972 1 Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly None

**ARCH RMNO** 01-05-03  
**RMTITLE** I.T. ROOM, 1ST FLOOR

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI 2603 0 Bench, Work for I.T. Technician Repair Area: Lights/Bins/Power/Data Strip floor 60"w x 24"d x Adj Ht E 115V x 2 Quad: 2 Data-com  
 OPVI 3062 1 Wastebasket, Office,Tall, 23 gal, with flip lid floor 20lx11wx30h None  
 OPVI 3972 1 Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly None

**ARCH RMNO** 01-05-04  
**RMTITLE** I.T. ROOM, 1ST FLOOR

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI 3062 1 Wastebasket, Office,Tall, 23 gal, with flip lid floor 20lx11wx30h None  
 OPVI 3972 1 Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly None

**ARCH RMNO** 02-05-10  
**RMTITLE** I.T. ROOM, 2ND FLOOR

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI 3062 1 Wastebasket, Office,Tall, 23 gal, with flip lid floor 20lx11wx30h None  
 OPVI 3972 1 Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly None

**DEPTNO**            **06**  
**DEPTTITLE**        **WINGS**  
**ARCH RMNO**        **01-06-01**  
**RMTITLE**            **OFFICE, WINGS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1005	3	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVICI	1500c	1	Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	2	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

**ARCH RMNO**        **01-06-02**  
**RMTITLE**            **OFFICE, WINGS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None

OPVI	0700	1 Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1005	3 Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1007	1 Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVICI	1500c	1 Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1 Robe Hook, Double	wall	6x6x3	
OPOI	3903	2 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

**DEPTNO** 07  
**DEPTTITLE** PHYSICAL THERAPY  
**ARCH RMNO** 01-07-00  
**RMTITLE** WAITING, PATIENT

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	3	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	6	Chair, Tandom, 1/2/3 seats w/various arm arnagements	floor	24x24x32	None
OPVI	1303	1	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	Varies per Space	None: Consider Backing
OPVI	1410a	1	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None

**ARCH RMNO** 01-07-01  
**RMTITLE** RECEPTION, PT

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	1	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-07-03**

**RMTITLE OFFICE, PT**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913	1	Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm

**ARCH RMNO 01-07-04**

**RMTITLE WORK SPACE, STAFF**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1500a	1	Workstation, Furniture System, With Transaction Top: Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	



OPOI	3903x	1	Computer, Personal, Complete with desktop Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
------	-------	---	--	---------	-----	------------------

**ARCH RMNO 01-07-05**

**RMTITLE PATIENT TOILET & DRESSING ROOM**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
CPCI	1805	1	Mirror, Full Length	wall	18x.5x60	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	9996	1	Bench, Sitting Suitable for Location	floor	Approx 48"w x 20"d x 20"h	None

**ARCH RMNO 01-07-06**

**RMTITLE PATIENT TOILET & DRESSING ROOM**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
CPCI	1805	1	Mirror, Full Length	wall	18x.5x60	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	9996	1	Bench, Sitting Suitable for Location	floor	Approx 48"w x 20"d x 20"h	None

**ARCH RMNO 01-07-07**

**RMTITLE TREATMENT, PATIENT**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	17x17x42	None
OPOI	3093	1	Container, Sharps 3 Gallon Wall Mount	wall	10.7"w x 6"d x 16.6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3904	1	Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPVI	7115b	1	Table, Treatment; EXISTING EQUIPMENT TO TX	floor	TBD	E 115V
OPVI	7118	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart: EXISTING TO TX	floor	20"w x15"d x48"h	E 115V

**ARCH RMNO 01-07-08**

**RMTITLE TREATMENT, PATIENT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	17x17x42	None
OPOI	3093	1	Container, Sharps 3 Gallon Wall Mount	wall	10.7"w x 6"d x 16.6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3904	1	Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPVI	7115b	1	Table, Treatment; EXISTING EQUIPMENT TO TX	floor	TBD	E 115V
OPVI	7118	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart: EXISTING TO TX	floor	20"w x15"d x48"h	E 115V

**ARCH RMNO 01-07-09**

**RMTITLE TREATMENT, PATIENT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None

OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	17x17x42	None
OPOI	3093	1	Container, Sharps 3 Gallon Wall Mount	wall	10.7"w x 6"d x 16.6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPVI	7115a	1	Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7118a	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart	floor	20"w x15"d x48"h	E 115V

**ARCH RMNO 01-07-10**  
**RMTITLE STORAGE, PT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3005	1	Cart, Linen, w/ 4 shelves and Staph Check enclosure	floor	36"w x 21"d x 64"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPVI	7108	1	Chilling Unit, Cold Packs, Complete w/packs, side table	floor	27"l x 16"d x 31"h	E 115; near drain
OPVI	7108a	1	Hydrocolator, Hot Pack, Unit: EXISTING EQUIPMENT TO TX	floor	27"l x 16"d x 33"h	E 115; near drain
OPVI	7108b	2	Rack for Hydrocolator towels/covers, folding hooks, 6/rack: REQUIRES BACKING	wall	36x15x2	None
CPCI	7122	1	Rack, Wall for Cane & Crutch, capacity of 6 crutch sets per support: REQUIRES BACKING	wall/backing	66"l x5.25"w x with hooks 9"d	None:Requires Backing
OPOI	9858a	1	Wheelchair: EXISTING EQUIPMENT TO TX	floor	24x30x42	None

**ARCH RMNO 01-07-11**  
**RMTITLE TREATMENT PATIENT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None

OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	17x17x42	None
OPOI	3093	1	Container, Sharps 3 Gallon Wall Mount	wall	10.7"w x 6"d x 16.6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPVI	7115a	1	Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7118a	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart	floor	20"w x15"d x48"h	E 115V

**ARCH RMNO 01-07-12**

**RMTITLE STORAGE, SOILED LINEN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	3016	2	Hamper, Linen, Mobile, with 2 Cloth Bag, Foot Operated Lid	floor	20 diam x 36	None

**ARCH RMNO 01-07-13**

**RMTITLE NOURISHMENT AREA**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPSVI	3310	1	Refrigerator, Undercounter, ADA 5.5 cubit ft, auto defrost, SS	floor	23.63"w x 23.5"d x 32"h	E 115V

**ARCH RMNO 01-07-14**

**RMTITLE EXERCISE AREA**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1006	5	Chair, Plastic, 4 legs, stackable	floor	20x21x34	None

OPVI	1335a	4 Stool, No Backrest, Adj. Ht. Casters, Foot ring	floor	15 diam	None
OPVI	2256a	1 Shelving Unit, Mobile, Adj Shelves, Bins, Bumper, Door Type TBD	floor	26"wx24"d x74"h	None
OPVI	2521	1 Board, Combination, Bulletin-Chalkless Marker 50/50	wall	36x1x24	None
OPOI	2944	1 Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	3 Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPSVI	3083	1 Station, Sanitizing Suitable for Gym Areas; Wipe & Hand Sanitizer Dispenser, Waste Container	floor	47" H x 12" W x 17" D	None
OPSVI	3102a	3 Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPCI	4136	3 Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data
OPCI	4138	1 Wall Mount for TV Monitors up to 55"; movement action TBD: OWNER TO SPECIFY: REQUIRES	wall	See Details	None
OPCI	4138a	2 Ceiling Mount for TV Monitors up to 55"; movement action TBD: REQUIRES BACKING	wall	See Details	None
OPOI	4144	1 Home Video Game Console (Wii station): EXISTING EQUIPMENT TBD?????	countertop	3"w x 9"d x 8"h	video cable connection
OPVI	7102a	1 Upper Body Exerciser-Step Thru Seating: SciFit Pro 1	floor	61"l x 30"w x 62"	E 115V; Data
OPVI	7103a	1 Bike, Exercise, Recumbent, Walk Thru Design, With Monitor; (For PT Area)	floor	67"l x 23"w x 54"h	E Ded 120V; 15 amp; Data
OPVI	7103c	1 Bike, Exercise, Recumbent; No step thru With Monitor: Precor (unit 219 lbs)	floor	63"l x23"w x54"h	E Ded 120V; 15 amp; Data
OPVI	7103d	2 Bike, Air: Assault	floor	51"l x23.5"w x35.5"h	None
OPOI	7107	1 Exercise Unit, Upper and Lower Body, BowFlex XLT Gym: EXISTING EQUIPMENT TO TX	floor	98lx50wx83h	None Req
OPOI	7107b	1 Exercise Unit, Total Gym, Fit: CLINICAL PKG	floor	93"l x 18.5"w x44.5"h	None
OPVI	7110a	1 Mirror, Glass, Tempered, 3 Section, Mobile: EXISTING EQUIPMENT TO TX	floor	30"w x10"d x72"h	None
OPVI	7110b	1 Storage, Multipurpose, for weights, includes rack, peg hooks: EXISTING TO TX	floor	24x18x72	None

OPVI	7113	1 Parallel Bars, Powered, Adult Adj Ht. & Width, 10' long: EXISTING EQUIPMENT TO TX	floor	up to 25"w x 10'l x42"h	E 120V-5 Amp: floor outlet
OPVI	7115a	1 Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7120a	1 Rack, Combination weight and dumbbells-opposite sides	floor	31"w x53"h x24"d	None
OPVI	7120b	1 Rack, Progression Weight, Static	floor	56"l x 23"d x 36"h	None
OPVI	7128b	1 Treadmill, Individual Walking Exercise, W/Handrails	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data- Com
OPVI	7128c	1 Treadmill, Individual Walking Exercise, W/Handrails: EXISTING EQUIPMENT TO TX	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data- Com
OPVI	7129a	1 Exercise Unit, Elliptical Trainer, w/progress monitor	floor	84"l x 31"w x67"	E 120V Data-Com
OPVI	7129b	1 Stepper, Recumbent: Sci Fit Step1; EXISTING EQUIPMENT TO TX	floor	68"l x30"w x48"h	E 115V; Data
OPVI	7135	1 Dynamometer, Hand, Hydraulic: EXISTING TO TX	stored	n/a	None
OPVI	7136	1 Biomechanical Ankle Platform System (BAPS): EXISTING TO TX	stored	n/a	None
OPVI	7137b	1 Rack for Storing Therapy Balls- Mobile	floor	64"l x 19"d x64"h	None
OPVI	7163	1 Body Solid Universal Gym	floor	12' w x 12'd x unknown ht	None
OPOI	8202	1 Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x12"	None

**DEPTNO**           **08**  
**DEPTTITLE**       **BEHAVIORAL HEALTH**  
**ARCH RMNO**       **01-08-00**  
**RMTITLE**           **WAITING, PATIENT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	2	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1010	2	Chair, Child	floor	18x18x24	None
OPVI	1019	7	Chair, Tandem, 1/2/3 seats w/various arm arrangements	floor	24x24x32	None
OPVI	1409	1	Table, Child Use, Square or Round (optional shape & colors)	floor	24" diameter 19"h	None
OPVI	1410a	2	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4132	1	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	38.1"w x 22.8"h x 4"d	E 115V; cabling; Data

**ARCH RMNO**       **01-08-01**  
**RMTITLE**           **RECEPTION, BEHAVIORAL HEALTH**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	

OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Comp
OPVI	3940	1	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-08-02**

**RMTITLE CIRCULATION (AED LOCATION) BETWEEN PC-EYE CLINICS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	8202	1	Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x 12"	None

**ARCH RMNO 01-08-03**

**RMTITLE OFFICE, THERAPIST**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary, Std.Back, 5-Legged, Ergonomic, Adj. Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase, Non-Sectional, Open, 3 Adj. Shelves	floor	36"w x 12"d x 40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"dx66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x 22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant, 28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x 24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	



OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
------	-------	---	--	---------	-----	------------------

**ARCH RMNO 01-08-05**

**RMTITLE OBSERVATION, PATIENT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	1	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPVI	2810	1	Vacuum, Hand, Portable, Rechargeable	stored	16"l x 9"w x 10"h	E 115V for charging
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-08-06**

**RMTITLE PLAY THERAPY (Panic Button)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0681	3	Cabinet Suitable for display of teaching materials or models, locking doors optional	floor	36"w x 18"d x 65"h	None
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1010	4	Chair, Child	floor	18x18x24	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1092	1	Rug, Area: Easy Clean up: Childrens Theming	floor	Varies	None

OPVI	1409	1 Table, Child Use, Square or Round (optional shape & colors)	floor	24" diameter 19"h	None
OPVI	1419	1 Table, Rectangular, On Casters, Optional Flip Top, Optional Drawer	floor	30"w x 24"d x 30"h	None
OPVI	2268	3 Shelves, Suitable for storage of toys, games, must be child accessible, secure to wall	floor/secu re	36"w x 18"d x 48"h	None
OPOI	2944	1 Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	1 Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None

**ARCH RMNO 01-08-07**

**RMTITLE OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary, Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	36"w x12"d x40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x24"dx66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO**            **01-08-08**  
**RMTITLE**            **OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	36"w x12"d x40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x24"dx66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO**            **01-08-09**  
**RMTITLE**            **OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	36"w x12"d x40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x24"dx66"h	none

OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-08-10**  
**RMTITLE GROUP THERAPY ROOM (Panic Button- DIMMABLE LIGHT)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0211	12	Chair, Conference Type, Optional On Casters, w/arms	floor	24.5x28x37	None
OPVI	0552	1	Credenza, w/4 doors, 2 drawers	floor	72x24x30	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1402a	4	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx d TBD x30	E 115V-floor: Data-Com
OPVI	2524	1	Board, Marker, Cleanable, OPTIONAL AS PORTABLE	wall	28x2x24	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPOI	2945	1	Telephone, Conference: TO BE SPECIFIED BY OWNER	desktop	14X13X3	E 115V; Data-Com
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None

OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216a	1	Coffeemaker, Low Profile; Pour Over Brewing w/ 2 Warmer Burners	countertop	8.75"w x17"d x16 7/8"h	E 120V Ded: 14 amps; Water Supply
OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4136	1	Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

**ARCH RMNO 01-08-11**

**RMTITLE OFFICE, MEDICAL SOCIAL WORKER (Panic Button)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary, Std.Back, 5-Legged, Ergonomic, Adj. Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant, 28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-08-12**

**RMTITLE OFFICE, APRN (Panic Button)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary, Std.Back, 5-Legged, Ergonomic, Adj. Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op

OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-08-13**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-08-14**

**RMTITLE OFFICE, THERAPIST, BEHAVIORAL HEALTH**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	36"w x12"d x40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x24"dx66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double	wall	6x6x3	
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-08-15**

**RMTITLE OFFICE, THERAPIST, BEHAVIORAL HEALTH**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	36"w x12"d x40"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x24"dx66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52x30x32	None
OPVI	1026	1	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPVI	9312a	1	Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer,	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps



**DEPTNO**            **09**  
**DEPTTITLE**        **PRIMARY CARE**  
**ARCH RMNO**        **01-09-00**  
**RMTITLE**            **RECEPTION, PRIMARY CARE: 5 STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	5	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	5	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPOI	2941	5	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	5	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3907	5	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3916	2	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Comp
OPOI	3936b	5	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	5	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO**        **01-09-02**  
**RMTITLE**            **ALCOVE, WEIGHT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1758	1	Casework, Modular, Designed to Meet Space Req: WEIGHT ALCOVE: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3041a	1	Wastebasket, Step-On, Plastic, Rectangular, Color TBD	floor	approx 16"w x12"d x23"h	None
OPVI	8911	1	Scale,Person Weighing, 1000 lb capacity, for wheelchair access, stadiometer	floor	28"w x30"dx 55"h	E 115V; or battery op

**ARCH RMNO**        **01-09-04**  
**RMTITLE**            **EXAM, PC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	1008	1 Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1 Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1 Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1 Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1 Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1 Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPVI	8572	1 Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer	stored	n/a	E 120V for charger or Battery Operated
OPOI	8776	1 Integrated DX Sys; Otoscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312a	1 Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer,	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps
OPVI	9854	1 Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None
OPVI	9991	1 Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None

CPCI 9993 1 Track, Cubicle Privacy Curtain, of ceiling as required-see layout None:  
Appropriate Size & fabric type

**ARCH RMNO 01-09-05**  
**RMTITLE EXAM, PC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPVI	8572	1	Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer	stored	n/a	E 120V for charger or Battery Operated
OPOI	8776	1	Integrated DX Sys; Otoscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312b	1	Table, Patient Exam, Power, w/IQ scales,Lowers to 14" w/clean assist roller sys/heated drawer/recep	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps

OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None
OPVI	9991	1	Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

**ARCH RMNO 01-09-06**  
**RMTITLE EXAM, PC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPVI	8572	1	Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer	stored	n/a	E 120V for charger or Battery Operated

OPOI	8776	1	Integrated DX Sys; Otoloscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312a	1	Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer,	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None
OPVI	9991	1	Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

**ARCH RMNO 01-09-07**

**RMTITLE MEDICATION ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	5018	1	Sink, Smart for drug waste containment	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPSVI	5030b	1	Controlled Dispensing System - 2 Drawer Main-Countertop Unit (LEASE ITEM?)	Countertop	23"W x27"D x29"H	E 115V, Data-Com x 2
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	30"w x 80"h x 29.5"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 31.5"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPVI	9299	1	Stand, IV, with 2 hangers, 5 legged, mobile	floor	20" diameter	None

## RMTITLE TREATMENT ROOM

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1335	1	Stool, Clinical, NO Backrest, Pneumatical Adjustable Base, Optional Footring	floor	17x adj ht	None
OPVI	1752	1	Casework, Modular, Designed to Meet Space Req: TYPICAL TREATMENT ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	8166	1	Cart, Procedures Set Up	floor	23"w x25"d x40"h	None
OPVI	8261	1	Electrosurgical Unit, (Hyfrecator) Minor Procedures - Cautery, on stand	floor	18"diam x60	E 115V
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPCI	8570b	1	Light, Minor Surgical,Ceiling Pedestal Mount Single Arm: CEILING SUPPPORT REQUIRED	ceiling	22 diam., 30-54	E,Struc.Support, 120VAC
OPSVI	8705	1	Monitor, Vital Signs; BP, Pulse, SPO2, Resp, EKG, On Optional Cart	on stand	14"w x 12"d x 12"h	E 115V; Data/Com
OPVI	8791	1	Oximeter, Pulse, Handheld/Portable	stored	n/c	battery operated

OPVI	8801	1	Pump, Infusion on Mobile Stand	floor	20diam x65	E 115V(recharging)
OPVI	9301	1	Stand, Mayo, Removable Tray,Adj.Ht.	floor	16x20x40	None
OPVI	9312c	1	Table, Universal Procedure,(lithotomy/trendelenberg/podiatry)	floor	28wx18-40hx72	E,115VAC, 10amp, 50/60Hz
OPVI	9991	1	Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

**ARCH RMNO 01-09-12**

**RMTITLE CLEAN ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1757	1	Casework, Modular, Designed to Meet Space Req: CLEAN ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	5274	1	Incubator, Dry Heat, Open, 10/20 well, digital	countertop	11.3"w x 7.9"d x 3.9"h	E 115V
OPVI	9221	1	Sterilizer, Table Top,Electric,Steam Type,w Printer & Pouch Rack: VOLTAGE TBD; SEE MFG DETAILS	countertop	23.8"w x17.8"d x17.8"h	E 115 OR 230 VAC, 50/60 Hz, 15 A Single Phase

**ARCH RMNO 01-09-13**

**RMTITLE SOILED ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1756	1	Casework, Modular, Designed to Meet Space Req: SOILED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3016	1	Hamper, Linen, Mobile, with 2 Cloth Bag, Foot Operated Lid	floor	20 diam x 36	None
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19"w x 16"d x 32"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None

OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

**ARCH RMNO 01-09-14**  
**RMTITLE STORAGE, CLEAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	6	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-09-15**  
**RMTITLE OFFICE, FLOATING FOR 8 PROVIDERS**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	8	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	8	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1303	1	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	Varies per Space	None: Consider Backing
OPVICI	1500g	1	Workstation,Open Office Furniture System,Modular, 8 staff	floor	See Design	E Quad x 8; data/com x4
OPOI	2941	8	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	8	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3903	8	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	8	Keyboard Tray, Style, Size TBD	under desk	TBD	None



**ARCH RMNO 01-09-16**  
**RMTITLE OFFICE, PC PHYSICIAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-09-17**  
**RMTITLE OFFICE, PC PHYSICIAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-09-18**  
**RMTITLE OPEN WORK, TEAM SPACE: 6 STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	0150	6 Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	6 Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1 Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1406	1 Table, Work Suitable for Function of Room	floor	72"lx36"d x42"h	None
OPVICI	1500f	1 Workstation,Open Office Furniture System,Modular, 6 staff	floor	See Design for BACKING REQUIREMENTS	E Quad x 6; data/com x4
OPOI	2941	6 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	6 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3903	6 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	6 Keyboard Tray, Style, Size TBD	under desk	TBD	None

**DEPTNO**            **10**  
**DEPTTITLE**        **EYE CARE**  
**ARCH RMNO**        **01-10-00**  
**RMTITLE**            **DISPENSING, EYEWEAR**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPOI	3904	2	Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPSVI	6400	1	Eye Wear Dispensing Furnishing Package by Fashion Optical: SEE FASHION OPTICAL DETAILS	floor	See Design	See Design
OPVI	6402	2	Chair, Optometric Dispensing Technician, w/Back,Casters	floor	24x24x30	None
OPVI	6402a	4	Chair, Optometric Dispensing, No arms, For Patient Use	floor	24x24x30	None

**ARCH RMNO**        **01-10-00a**  
**RMTITLE**            **WORK AREA, TECHNICIAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0182	1	Chair, Rotary,Std.Back,4-Legs, Nesting,Ergonomic, Partial Arm, Standing Height	floor	24.5x28x48	None
OPVI	1202	2	Lighting, Task, Overhead, LED, Warm Light (electrical behind upper cabinets)	under overhead s	TBD by Space	E 115V Receptacle Behind Uppers
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	6412	1	Safe, Small, for storing moneys; EXISTING EQUIPMENT TO TX		TBD	None

OPVI	6450	1	Warming Unit,for Frame Adjustment, Hot Air Warmed	desktop	6.5"w x12"l x6"h	E 110V
------	------	---	---	---------	------------------	--------

**ARCH RMNO 01-10-02**  
**RMTITLE TESTING, PRELIMINARY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	0	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3902	1	Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3904	1	Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPVI	6411	1	Cart, Supply, Portable, Soft Edge with Handle Top; Optional locking	floor	18"w x 18"d x33"h	None
OPVI	6420	1	Ultrasound, B Scan, handheld	stored	n/a	Battery Operated
OPVI	6421	1	Pachymeter, System, blue tooth technology, printer (EXISTING EQUIPMENT TO TX)	countertop	printer 4"w x 6"d x1.54"h	E 115 for printer & charging
OPVI	6422	1	Refractor/Keratometer/Topographer, Automated; Sits on a table top #6439	tabletop	N/A	E-Ded: Data
OPVI	6437	1	Stool, Backrest,Adj.Ht.,Footring,Casters	floor	18x18x36	None
OPVI	6439	1	Tables, Instrument,Casters, Motorized, height adjustable,	floor; mobile	36"wx 20"d x adj ht	E-Dedicated, 1 Data
OPVI	6444b	1	Visual Field Analyzer, Automated Humphrey Matrix 800: EXISTING EQUIPMENT TO TX	shared table to	12"w x 22"d x17"h	E-115V Dedicated; Data
OPVI	6446	2	Tonometer, Hand Held	stored	n/a	None
OPVI	6489	1	Lensometer, Computerized	countertop	6.3"w x 16.53"d x12.2"h	E 115; Data
OPVI	6501	1	Monitor, Blood Pressure on stand (EXISTING EQUIPMENT TO TX)	floor	18" diameter	E 115V for charging
OPVI	6509	1	Tester, Brightness Acuity	stored	n/a	None

OPVI	6510	1	Tests, Optometric, Various: EXISTING EQUIPMENT TO TRANSFER	stored	n/a	None
OPVI	6528	1	Test, Color Vision: HHR standard version/with illuminator (EXISTING EQUIPMENT TO TX)	stored	6"w x 5"h x .5"d	None

**ARCH RMNO 01-10-03**  
**RMTITLE EXAM, EYE (RM W/CHAIR MOVER)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step- On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1	Chair, Eye Exam, Powered	floor	86x20/56x65	E, Lighting Control
OPVI	6406a	1	Floor Unit, Stand & Instrument Control	floor	24x24x65	E-115V
OPVI	6406c	1	Chair Glide, Mover, To Accommodate Wheelchairs, for bases up to 22.5" wide	under chair	varies	None
OPVI	6417	3	Handles, Rechargable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1	Refractor for Stand	attached	n/a	None
OPVI	6425	1	Ophthalmoscope,Indirect,Binocular	stored	n/a	Battery Operated

OPVI	6425a	1 Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1 Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPVI	6429a	1 Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data
OPVI	6430	1 Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1 Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1 Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand
OPVI	6437	2 Stool, Backrest,Adj.Ht.,Footring,Casters	floor	18x18x36	None
OPVI	6445	1 Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1 Tonometer Mount	attached	N/A	None
OPVI	6448	1 Frame,Trial Lens,Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None
OPVI	6449	1 Trial Lens Set,Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPVI	6452	1 Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software:Wall Mount: BACKING	wall	TBD	E 115V: Data-Com
OPVI	6481	1 Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1 Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1 Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1 Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1 Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1 Forceps, Jeweler	stored	n/a	None
OPVI	9127	1 Sphygmomanometer,Aneroid,Hand Use,Portable, With Assortment	stored	n/a	None

**ARCH RMNO**            **01-10-04**  
**RMTITLE**              **EXAM, EYE**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1	Chair, Eye Exam, Powered	floor	86x20/56x65	E, Lighting Control
OPVI	6406a	1	Floor Unit, Stand & Instrument Control	floor	24x24x65	E-115V
OPVI	6417	3	Handles, Rechargeable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1	Refractor for Stand	attached	n/a	None
OPVI	6425	1	Ophthalmoscope,Indirect,Binocular	stored	n/a	Battery Operated
OPVI	6425a	1	Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1	Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPVI	6429a	1	Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data

OPVI	6430	1	Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1	Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1	Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand
OPVI	6437	2	Stool, Backrest,Adj.Ht.,Footring,Casters	floor	18x18x36	None
OPVI	6445	1	Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1	Tonometer Mount	attached	N/A	None
OPVI	6448	1	Frame,Trial Lens,Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None
OPVI	6449	1	Trial Lens Set,Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPVI	6452	1	Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software:Wall Mount: BACKING	wall	TBD	E 115V: Data-Com
OPVI	6481	1	Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1	Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1	Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1	Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1	Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1	Forceps, Jeweler	stored	n/a	None
OPVI	9127	1	Sphygmomanometer,Aneroid,Hand Use,Portable, With Assortment	stored	n/a	None

**ARCH RMNO 01-10-05**  
**RMTITLE SPECIAL TESTING**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None



OPVI	1015	1	Chair, Patient: Consider Sturdy Wide Base On Casters for Patient Convenience?	floor	22"w x 20"d x std height	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3902	1	Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3904	1	Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPVI	6403	1	Ocular Coherence Tomography, with monitor, software & table (EXISTING EQUIPMENT TO TX)	tabletop	table: 39"l x 22"w x adj ht	E 115V-DED, Data-Com
OPVI	6404	1	Camera, Fundal/Pt Mgmt Software System/on table (EXISTING EQUIPMENT TO TX)	sits on table	table: 36"w x 20"d x adj ht	E 115V,-Dedicated, Data-Com
OPVI	6437	2	Stool, Backrest,Adj.Ht.,Footring,Casters	floor	18x18x36	None
OPVI	6444	1	Visual Field Analyzer, Automated; SITS ON ITS OWN TABLE: EXISTING EQUIPMENT TO TX	tabletop	36"w x24"d x ADJ HT	E 115V-Dedicated; Data

**ARCH RMNO 01-10-06**

**RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

**ARCH RMNO 01-10-07**

**RMTITLE EXAM, EYE (RM W/CHAIR MOVER)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None

OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1 Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1 Chair, Eye Exam, Powered	floor	86x20/56x65	E, Lighting Control
OPVI	6406a	1 Floor Unit, Stand & Instrument Control	floor	24x24x65	E-115V
OPVI	6406c	1 Chair Glide, Mover, To Accommodate Wheelchairs, for bases up to 22.5" wide	under chair	varies	None
OPVI	6417	3 Handles, Rechargable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1 Refractor for Stand	attached	n/a	None
OPVI	6425	1 Ophthalmoscope,Indirect,Binocular	stored	n/a	Battery Operated
OPVI	6425a	1 Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1 Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPVI	6429a	1 Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data
OPVI	6430	1 Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1 Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1 Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand
OPVI	6437	2 Stool, Backrest,Adj.Ht.,Footring,Casters	floor	18x18x36	None
OPVI	6445	1 Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1 Tonometer Mount	attached	N/A	None

OPVI	6448	1	Frame,Trial Lens,Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None
OPVI	6449	1	Trial Lens Set,Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPVI	6452	1	Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software:Wall Mount: BACKING	wall	TBD	E 115V: Data-Com
OPVI	6481	1	Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1	Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1	Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1	Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1	Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1	Forceps, Jeweler	stored	n/a	None
OPVI	9127	1	Sphygmomanometer,Aneroid,Hand Use,Portable, With Assortment	stored	n/a	None

**ARCH RMNO 01-10-08**  
**RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

**ARCH RMNO 01-10-09**  
**RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

**ARCH RMNO 01-10-10**  
**RMTITLE OFFICE, PROVIDER**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	36x12x65	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-10-11**  
**RMTITLE OFFICE, PROVIDER**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	36x12x65	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-10-12**  
**RMTITLE OPEN WORK, TEAM SPACE (4 STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	0150	4 Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	4 Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1 Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1406	1 Table, Work Suitable for Function of Room	floor	72"lx36"d x42"h	None
OPVICI	1500e	1 Workstation,Open Office Furniture System,Modular, 4 staff, including uppers	floor	See Design	E Quad x 4; data/com x4
OPOI	2941	4 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3061	1 Wastebasket, Office, 16 Gallon Slim Jim w/ Venting Channels Pack of 4	floor	22"w x11"l x25"h	None
OPOI	3902	1 Workstation On Wheels; incldg monitor/keyboard/computer	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3904	3 Computer, Laptop, wireless connectivity; OPTIONAL-portable cart	attached	TBD	E 115V; Data-Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

**DEPTNO** 11  
**DEPTTITLE** WIC  
**ARCH RMNO** 01-11-02  
**RMTITLE** BREASTFEEDING ROOM (REQ LIGHT DIMMER)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	TBD	Battery Op
OPVI	1160	1	Recliner, Patient, with Glider Option; Wall Saver, Clean Out,	floor	28.5"w x 39.75"d x 49.5"h	None
OPVI	1205	1	Lamp, Table, Decorator, 3 level bulb	desktop	15 diam x 30	E 115V
OPVI	1410a	1	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	1755	1	Casework, Modular, Designed to Meet Space Req: BREAST FEEDING ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	2294	1	Diaper Pail, with Step	floor	12"w x 10"d x 32H	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
CPCI	3135	1	Diaper Changing Station, Horizontal Drop Down	wall	22"d x35.5"l x5.25"h	None

**ARCH RMNO** 01-11-03  
**RMTITLE** OFFICE, WIC

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none

OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

**ARCH RMNO**      **01-11-04**  
**RMTITLE**        **OFFICE, WIC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

**ARCH RMNO**      **01-11-05**  
**RMTITLE**        **STORAGE, WIC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None

OPVI	2254	6	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO**            **01-11-06**  
**RMTITLE**            **OFFICE, WIC**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

**ARCH RMNO**            **01-11-07**  
**RMTITLE**            **OFFICE, WIC**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com



OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Comp

**ARCH RMNO**            **01-11-08**  
**RMTITLE**                **EDUCATION, HEALTH**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	1005	3	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVICI	1500c	1	Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	2	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	2	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**DEPTNO** 12  
**DEPTTITLE** SPECIALTY CARE  
**ARCH RMNO** 01-12-00  
**RMTITLE** MEDICATION ROOM (FOR PHN/WIC/AND OTHERS)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	5018	1	Sink, Smart for drug waste containment	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPSVI	5030b	1	Controlled Dispensing System - 2 Drawer Main-Countertop Unit (LEASE ITEM?)	Countertop	23"W x27"D x29"H	E 115V, Data-Com x 2
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	30"w x 80"h x 29.5"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 31.5"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data

**ARCH RMNO** 01-12-01  
**RMTITLE** EXAM, SPECIALTY

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None



OPVI	1008	1 Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1 Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1 Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPVI	1750	1 Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1 Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1 Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1 Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPVI	8572	1 Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer	stored	n/a	E 120V for charger or Battery Operated
OPOI	8776	1 Integrated DX Sys; Otoscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312a	1 Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer,	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps
OPVI	9854	1 Sundry Jar Set of 5, for Exam Room Supplis (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None

OPVI	9991	1	Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

**ARCH RMNO 01-12-03**  
**RMTITLE EXAM, SPECIALTY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8x8x10	None
OPVI	8572	1	Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer	stored	n/a	E 120V for charger or Battery Operated

OPOI	8776	1	Integrated DX Sys; Otoloscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312a	1	Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer,	floor	27wx18-37hx72	E 115VAC,50/60Hz, 8amps
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None
OPVI	9991	1	Chart Rack, HIPPA Application: BACKING REQUIRED: Typically outside of exam rooms	wall	15x3x12	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

**ARCH RMNO 01-12-04**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	5	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-12-05**

**RMTITLE OFFICE, SPECIALTY PHYSICIAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-12-06**  
**RMTITLE OFFICE, SPECIALTY PHYSICIAN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-12-07**  
**RMTITLE OPEN WORK, TEAM SPACE (6 STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	6	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	6	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1406	1	Table, Work Suitable for Function of Room	floor	72"lx36"d x42"h	None
OPVICI	1500f	1	Workstation,Open Office Furniture System,Modular, 6 staff	floor	See Design for BACKING REQUIREMENTS	E Quad x 6; data/com x4
OPOI	2941	6	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	6	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3903	6	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

OPVI	3940	6 Keyboard Tray, Style, Size TBD	under desk	TBD	None
------	------	----------------------------------	------------	-----	------



**DEPTNO** 13  
**DEPTTITLE** PUBLIC HEALTH  
**ARCH RMNO** 01-13-00  
**RMTITLE** RECEPTION, SHARED WITH WIC

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0152	1	Chair, Rotary, Bariatric, Std.Back, On Casters, Adjustable Arms	floor	TBD	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3907x	1	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPOI	3936a	1	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	1	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO** 01-13-02  
**RMTITLE** ALCOVE, WEIGHT

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1758	1	Casework, Modular, Designed to Meet Space Req: WEIGHT ALCOVE: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None

OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVI	8915	1	Scale,Multipurpose, Stand On, Low Profile, stadiometer	floor	24"w x26"d 46"h	E 115V; or Batter Operated
OPVI	9310	1	Table, Pediatric Exam, Digital Weight with Measurement, Storage underneath	floor	42x26x52	E115V;

**ARCH RMNO 01-13-04**  
**RMTITLE OFFICE, INFECTION CONTROL**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914	1	Printer, Multifunction, LaserJet mf copy/scan/fax	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

**ARCH RMNO 01-13-05**  
**RMTITLE EXAM, PHN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8572	1	Light, Examination,Halogen, Wall/Board or Table Mount	wall or table	6x6x30	E 115V
OPOI	8776	1	Integrated DX Sys; Otoscope-Ophthalmoscope-BP-Optional Therm; probe cover disp:	wall mount	30"w x 16"h x 6"d	E 115V
OPVI	9312x	1	Table, Patient Examination, Powered EXISTING EQUIPMENT TO TRANSFER	floor	TBD	E 115VAC
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None

**ARCH RMNO 01-13-06**  
**RMTITLE PHN MEDICATION ADMINISTRATION (MIDMARK CABINETRY?)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Locking Casters	floor	17x17x42	None
OPVI	1419	1	Table, Rectangular, On Casters, Optional Flip Top, Optional Drawer	floor	30"w x 24"d x 30"h	None
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16x16x18	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: finish options to be selected	wall mount	24"w x 12"h x 6"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	30"w x 80"h x 29.5"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 31.5"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	stored	varies	None

**ARCH RMNO**            **01-13-07**  
**RMTITLE**              **OFFICE, PHN**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	

OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
------	-------	---	--	---------	-----	------------------

**ARCH RMNO 01-13-08**

**RMTITLE OPEN WORK AREA, PHN (? # STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	3	Chair, Rotary, Std.Back, 5-Legged, Ergonomic, Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	3	Cabinet, Filing, Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVICI	1500d	1	Workstation, Open Office Furniture System, Modular, 3 person	floor	Various	E Quad x 3: Data Com x 3
OPOI	2941	3	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3	Wastebasket, Rectangular, plastic, water/fire resistant, 28qt	floor	14x11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	3	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3940	3	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 01-13-09**

**RMTITLE ALCOVE, COPY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	3062	1	Wastebasket, Office, Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm

**ARCH RMNO 01-13-10**

**RMTITLE STORAGE, PHN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0680	2	Cabinet, Storage, 6 Adj.Shelves, 2 Doors, Lock	floor	36x18x65	None
OPVI	2254	4	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts, wdts, depths	floor	Various	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None

OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
------	------	---	---	-------	-------------	------

**ARCH RMNO 01-13-11**  
**RMTITLE STORAGE, PHN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	2254	3	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None
------	------	---	--	-------	---------	------

OPVI	2403	2	Cart, Stainless Steel, 3 Shelf, Heavy Duty, 1000 lb capacity	floor	25.75"w x 48"l x 37 3/8"h	None
------	------	---	--	-------	---------------------------	------

OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
------	------	---	---	-------	-------------	------

**ARCH RMNO 01-13-12**  
**RMTITLE OPEN WORK AREA (4 STAFF)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	0150	4	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
------	------	---	--	-------	------------	------

OPVI	0430d	4	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
------	-------	---	--	-------	-----	------

OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
------	-------	---	---	-------	-----	------

OPVICI	1500e	1	Workstation,Open Office Furniture System,Modular, 4 staff, including uppers	floor	See Design	E Quad x 4; data/com x4
--------	-------	---	---	-------	------------	-------------------------

OPOI	2941	2	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
------	------	---	---	---------	------------------	----------

OPOI	2941x	2	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
------	-------	---	---	---------	------------------	----------

OPVI	3039a	4	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
------	-------	---	--	-------	-------------	------

OPOI	3903	2	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
------	------	---	--	---------	--------------------	------------------

OPOI	3903x	2	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
------	-------	---	--	---------	-----	------------------

OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
------	------	---	--	------------	------------------------	-----------------

OPVI	3940	4	Keyboard Tray, Style, Size TBD	under desk	TBD	None
------	------	---	--------------------------------	------------	-----	------

**DEPTNO** 14  
**DEPTTITLE** COMMUNITY SPACES  
**ARCH RMNO** 01-05-02  
**RMTITLE** STORAGE, KITCHEN (PREVIOUSLY TANK STORAGE)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None

**ARCH RMNO** 01-14-01  
**RMTITLE** CONFERENCE ROOM; LARGER OF THE 2 MTG SPACES

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0212	40	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	TBD	None
OPVI	1402a	20	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx d TBD x30	E 115V-floor: Data-Com
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4112	1	Projector, Ceiling Mount, wireless capability TO BE SPECIFIED BY OWNER: STRUCTURAL	ceiling	13"w x 10.5"d x 4.8"h	E 115V; Data-Com
OPCI	4120b	1	Screen, Projection, Powered, HDTV Format; Concealed in Ceiling: STRUCTURAL CONSIDERATIONS	wall/ceiling	139"w x8"d x78"h	E 120V, Control Switch
OPVI	4123	1	Board, Smart; Portable on Casters	FLOOR	APPROX 60" WIDE	None
OPVI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x5.12"h x2.6"d	POE
OPOI	8202	1	Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x12"	None

**ARCH RMNO** 01-14-02  
**RMTITLE** CONFERENCE ROOM (SMALLER OF THE 2 MTG SPACES)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0212	30	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	TBD	None
OPVI	1402a	15	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx d TBD x30	E 115V-floor: Data-Com
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None

OPCI	4112	1	Projector, Ceiling Mount, wireless capability TO BE SPECIFIED BY OWNER: STRUCTURAL	ceiling	13"w x 10.5"d x 4.8"h	E 115V; Data-Com
OPCI	4120b	1	Screen, Projection, Powered, HDTV Format; Concealed in Ceiling: STRUCTURAL CONSIDERATIONS	wall/ceiling	139"w x 8"d x 78"h	E 120V, Control Switch
OPVI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x 5.12"h x 2.6"d	POE
OPVI	4151	1	Lectern, For Education Environments	floor	30"w x 42"h x 24"d	None

**ARCH RMNO 01-14-04**

**RMTITLE KITCHEN, LIGHT COMMERCIAL, WARMING KITCHEN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3047a	1	Trash Collector w/Lid, and Dolly, 44 gallons (color TBD by Function)	floor	21.5" diameter x 36"	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x 4"d x 11.5"h	None
CPCI	3159	1	Sink, Handwashing, w/Eye Wash station	wall	10"d x 15"w x 5"h	H/C water; drain
CPCI	3160	1	Sink, 3 Well, SS, Drop In, Deep Bowl, Goose Neck Faucet	drop in sink	56"w x 25"d x 12"h	H & C Water, Drain
CPCI	3192	1	Hood Exhaust with Required Fire Suppression for cooking areas; optional CFM	wall mount	TBD	TBD
OPVI	3215	1	Coffeemaker, Automatic, Air Pot Twin Brewing System	countertop	17.75"w x 17.25"d x 22"h	E 120/240V Dedicated; 4 Wire; 17 amp; Water Supply
OPVI	3216b	1	Rack, Coffee Container for Air Pots, with drip trays	stored	23"w x 25"d x 15"h	None
OPVI	3216c	1	Pots, Air; Glass lined; 2.2 liter; lever action; hot up to 8 hrs (pkg 6 qty per case)	stored	brew ht 12.25"h/lid closed 14 9/16" h	None
OPCI	3250	1	Ice Station, Floor Model, Bulk: With Optional Hotel Ice Dispensary	floor	22"w x 31"d x 75.5"h	See Mfg Details
OPVI	3255	2	Microwave Oven, Commercial	overshelf	22"w x 19"d x 14"h	E 120V: 15 amp: NEMA 5-15
OPCI	3266	1	Range, Free Standing, Electric, Dual Oven Compartments	floor	29 15/16"w x 26 1/8"d x 47 1/8"h	E Ded 120-240V; 60Hz; 40-50 Amp
CPCI	3268	1	Refrigerator, Freezer: Commercial, SS, 2 Doors, on casters, 6 shelves, 1HP, exterior temp display	floor	54 1/8"w x 29.5"d x 84"h	E Ded 115V; 9.6 amps; NEMA 5-15P



OPCI	3352b	1 Dishwasher, Automatic, Undercounter, Commercial Unit; Reaches 185 degree water temp on	floor	33"h x 24"w x 24"d; See Details	E 3AC-208V;3x20 Amp;NEMA L15-20;1 Hot,2 Cold;drain
------	-------	--	-------	------------------------------------	--

**DEPTNO** 15  
**DEPTTITLE** PHARMACY  
**ARCH RMNO** 01-15-00  
**RMTITLE** WAITING, PHARMACY: (SMALL SEATING GROUPS)

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	7	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	22	Chair, Tandom, 1/2/3 seats w/various arm arnagements	floor	24x24x32	None
OPVI	1024	2	Seating, 3 Person, Couch Type	floor	84x30x28	None
OPVI	1026	4	Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1302	2	Rack, Magazine/Pamphlets,Wall Mounted 1 Pocket	wall	15.5" w x4"d x30"h	None: Consider Backing
OPVI	1410a	11	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4132	2	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	38.1"w x 22.8"h x 4"d	E 115V; cabling; Data
OPVI	5010	1	Medication Disposal System; Drop Box; 18 or 38 Gal; for Schedules II-V; Secures to Wall or Floor	floor	22.5"w x 22.5"l x 43.75"h	None
OPSVI	5048	1	Kiosk, Script Pro: FUTURE CONSIDERATION (for waiting area)	TBD	TBD	E 115V Ded; Data-Com
OPSVI	5049	2	Monitor, Status Message System: Bingo Boards, Script Pro: BACKING REQUIRED	TBD	TBD	E 115V Ded; Data-Com

**ARCH RMNO** 01-15-01  
**RMTITLE** RECEPTION, PHARMACY

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0150x	2	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	24.5x28x37	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPOI	2941x	2 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPOI	3903	3 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3940	3 Keyboard Tray, Style, Size TBD	under desk	TBD	None
OPSVI	5044	1 Signature Pad, Electronic: Integrated PC via USB port; Specified by Owner	countertop	7.5"w x 4.75"h x7.5"d	None
OPSVI	5044x	2 Signature Pad, Electronic: Integrated PC via USB port; EXISTING EQUIPMENT TO TX	countertop	7.5"w x 4.75"h x7.5"d	None

**ARCH RMNO**            **01-15-02**  
**RMTITLE**                **OFFICE, CONSULT**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO 01-15-03**  
**RMTITLE OFFICE, CONSULT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO 01-15-04**  
**RMTITLE OFFICE, HC CONSULTANT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None

CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO 01-15-05**  
**RMTITLE OFFICE, CONSULT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO 01-15-06**  
**RMTITLE OFFICE, CONSULT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	0 Robe Hook, Double	wall	6x6x3	
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1 Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO**            **01-15-07**  
**RMTITLE**            **OFFICE, CONSULT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPVI	0382	0	Bookcase,Non-Sectional,Open, 2 Adj. Shelves	floor	36x12x30	NONE
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	0	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;data/com

**ARCH RMNO**            **01-15-08**  
**RMTITLE**            **BREAK ROOM, STAFF \***

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1005	2	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None

OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPVI	3216d	1	Coffeemaker, Drip Through-Pour Over, 10 cup, always ready, 1 burner	countertop	8x12x14	E 115V
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	33"w-doors open 36.5" x 33 3/8"d x 70"h	E 115V; Water Supply for Icemaker
OPVI	3349	1	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.75"h x24"w x 19.43"d	E 120V: 13.9 amp

**ARCH RMNO 01-15-10**

**RMTITLE STORAGE, RX \***

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPVI	2254	4	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 01-15-11**

**RMTITLE BREAK OUT**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPVI	2407	1	Cart, Stainless Steel, 3 Shelf, On Casters, With Ergonomic Handle, 300 lb capacity	floor	16 1/2 "w x 28"; x 44 3/8"h w/handle	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
CPCI	3900	1	Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com

**ARCH RMNO 01-15-12**

**RMTITLE OFFICE, SUPERVISOR \***

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150x	1	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None

OPVI	0700x	1	Bookcase/Lateral file combination unit: EXISTING EQUIPMENT TO TX	floor	36x18x64	none
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-15-13**  
**RMTITLE STORAGE, RX**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

**ARCH RMNO 01-15-14**  
**RMTITLE PHARMACY AREA, MAIN**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150x	6	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	24.5x28x37	None
OPVI	0181x	6	Chair, Rotary,Std.Back:Extended Height: EXISTING EQUIPMENT TO TRANSFER	floor	24.5x28x48	None
OPSVI	1470	3	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPSVI	2273	6	Shelving,Pharmacy Storage; DESIGN BY RC SMITH; TYPE & QTY TBD	floor	SEE DESIGN DETAILS	NONE
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPOI	2941x	5	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	5	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062a	2	Wastebasket, Office,Tall, 23 gal, with flip lid (Blue Recycling-HIPPA)	floor	20lx11wx30h	None



OPVI	3062x	4	Wastebasket, Office,Tall, 23 gal, with flip lid (Blue Recycling-HIPPA): EXISTING EQUIPMENT TO TX	floor	20lx11wx30h	None
CPCI	3125	1	Rack, Coat, Size for up to 12 hooks for staff	wall	TBD	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm
CPCI	3900	6	Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com
OPOI	3903	5	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3903x	2	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913x	1	Printer, Laser, Individual EXISTING EQUIPMENT TO TX	countertop	TBD	E 115V; Data Comm
OPCI	5006	1	Filtered Water System, Digital, Fillmaster with 4 satellite stations	wall	6x6x12	E 115V; Water Supply
OPSVI	5007	8	System, Prescription Bag Hanging; bags, dividers; PART OF RC SMITH PKG	wall mounted	48"w x 48"h	None
OPOI	5018x	4	Sink, Smart for drug waste containment: EXISTING EQUIPMENT TO TX	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPOI	5021	2	Counter, Tablet-Kirby Lester; with built in scanning: EXISTING EQUIPMENT TO TRANSFER	countertop	8"w x 12.5"d x14.6"	E 110V
OPSVI	5027	1	Pharmacy, Robotic Medication Dispensing, EXISTING TO BE TRANSFERRED & COORDINATED	floor	152x35x76.5	See Scriptpro Design
OPOI	5043	1	Printer, Thermal, Label; (sits on ScriptPro System) EXISTING EQUIPMENT TO TX:	countertop	5.5"d x 14.5"h x 12"w	E115V;50/60hz;2amp;data-com
OPOI	5043x	1	Printer, Thermal, Label; (sits on ScriptPro System) EXISTING EQUIPMENT TO TX:	countertop	5.5"d x 14.5"h x 12"w	E115V;50/60hz;2amp;data-com
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	30"w x 80"h x 29.5"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5051x	1	Refrigerator, Scientific;Double door; EXISTING EQUIPMENT TO TX	floor	60"w x 80"h x 30"d	E-Ded 115V; 11.5 FLA;data
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;-15 to -30 Deg C;Onboard monitoring sys	floor	24"w x 31.5"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPSVI	5055	1	Refrigerator, Undercounter:ADA:Scientific;2-10 deg C;Onboard monitoring sys	floor	24"w x 31.5"h x 26.75"d	E-Ded 115V; 5 FLA;data

**ARCH RMNO** 01-15-15  
**RMTITLE** WAITING, PHARMACY (INTEGRATED EXISTING SPACE)

**GROUP EQNO QTY DESCRIPTION LOC SIZE UTILITIES**

OPVI	1007	2 Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	2 Chair, Tandom, 1/2/3 seats w/various arm arrnagements	floor	24x24x32	None
OPVI	1026	4 Seating, 1 Person, Upholstered: Style & Design TBD	floor	28x30x32	None
OPVI	1410a	2 Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	1421	1 Table, Sofa Suitable for Lounge/Waiting Areas	floor	36" diameter x 24"h	None
OPSVI	5049x	1 Monitor, Status Message System: Bingo Boards, Script Pro: BACKING REQ;EXISTING EQUIPMENT TO TX	TBD	TBD	E 115V Ded; Data-Com

**DEPTNO** 16  
**DEPTTITLE** ADMINISTRATION  
**ARCH RMNO** 01-16-00  
**RMTITLE** WAITING AREA, ADMINISTRATION

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	2	Chair, Tandom, 1/2/3 seats w/various arm arnagements	floor	24x24x32	None
OPVI	1410a	1	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4132	1	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	38.1"w x 22.8"h x 4"d	E 115V; cabling; Data

**ARCH RMNO** 01-16-01  
**RMTITLE** RECEPTION, ADMINISTRATION

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0151	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms, High Back	floor	24.5x28x37	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO** 01-16-02  
**RMTITLE** COPY/WORK ROOM

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
-------	------	-----	-------------	-----	------	-----------

OPVI	1405	2	Mail Cubbies (Risers) Optional 24 or 36 mail slots (REQUIRES BACKING)	wall mounted	TBD	None
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x36"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm
OPVI	3780	1	Meter, Postage, Automatic: High Productivity Unit;EXISTING EQUIPMENT TO PLACE ON	countertop	65"l x 25"d x 26"h	E 120V; Data/Com x 2
CPCI	3900	1	Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com

**ARCH RMNO 01-16-04**  
**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0680	2	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None

**ARCH RMNO 01-16-05**  
**RMTITLE OFFICE, ADMINISTRATOR (PANIC BUTTON)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	4	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1412	1	Table, Round, Diameter, 36"	floor	36 diamx30	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None

CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPCI	4131	1	Television, Flat Panel, 65"; LED; Smart TV; with mount; owner to coordinate: REQUIRES BACKING	wall/ceiling mt	57.6"w x 2.5"d x 33.2"h	E 115V; Cabling; Data

**ARCH RMNO**            **01-16-06**  
**RMTITLE**                **KITCHENETTE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1005	4	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1402a	1	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx d TBD x30	E 115V-floor: Data-Com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp; Water Supply; NEMA 5-5P
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	33"w-doors open 36.5" x 33 3/8"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	1	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.75"h x24"w x 19.43"d	E 120V: 13.9 amp

**ARCH RMNO**            **01-16-07**  
**RMTITLE**                **CONFERENCE ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0211	12	Chair, Conference Type, Optional On Casters, w/arms	floor	24.5x28x37	None
OPVI	1402	1	Table, Formal Conference Shape & Preferred Base Type TBD; (phone & computers)	floor	TBD	E 115V-floor; Data-Com
OPOI	2945	1	Telephone, Conference: TO BE SPECIFIED BY OWNER	desktop	14X13X3	E 115V; Data-Com
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None

OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3310	1	Refrigerator, Undercounter, ADA 5.5 cubit ft, auto defrost, SS	floor	23.63"w x 23.5"d x 32"h	E 115V
OPCI	4136	1	Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data
OPVI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x5.12"h x2.6"d	POE

**ARCH RMNO 01-16-08**  
**RMTITLE OFFICE, STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-16-09**  
**RMTITLE OFFICE, STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none

OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 01-16-10**  
**RMTITLE OFFICE, STAFF**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

**ARCH RMNO 01-16-12**  
**RMTITLE OFFICE, SECURITY (BY PRIMARY CARE)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	1005	2	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	24x24x34	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1 Robe Hook, Double	wall	6x6x3	
OPOI	3903x	1 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913	1 Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm
OPCI	4136	2 Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data



**DEPTNO** 17  
**DEPTTITLE** DENTAL  
**ARCH RMNO** 02-17-00  
**RMTITLE** WAITING, DENTAL

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1007	9	Chair, Patient, 4 legs, with arms, Bariatric	floor	31"w x22"d x 33"h	None
OPVI	1019	28	Chair, Tandom, 1/2/3 seats w/various arm arnagements	floor	24x24x32	None
OPVI	1302a	2	Rack, Magazine/Pamphlets,Wall Mounted 3 Pocket	wall	36x3x24	None: Consider Backing
OPVI	1410a	12	Table, End, Drum Round 18" Diameter	floor	18" Diameter x20"h	None
OPVI	3043	3	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPCI	4136	2	Television, Flat Panel, 55" Mount (wall or ceiling mount TBD) LED: OWNER TO SPECIFY: REQ	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

**ARCH RMNO** 02-17-01  
**RMTITLE** RECEPTION, DENTAL

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	3	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	TBD	None
OPVI	1480	3	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941	3	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	3760	1	Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3903	3	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

OPOI	3936a	3	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	3	Keyboard Tray, Style, Size TBD	under desk	TBD	None

**ARCH RMNO 02-17-04**

**RMTITLE CONFERENCE ROOM-BREAK ROOM**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0180	24	Chair, Rotary, Std.Back,4-Legs, Nesting,Ergonomic,NO Arms	floor	24.5x28x37	None
OPVI	1402a	4	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx d TBD x30	E 115V-floor: Data-Com
OPVI	2524	1	Board, Marker, Cleanable, OPTIONAL AS PORTABLE	wall	28x2x24	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal	floor	18"wx32"hx12"D	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp; Water Supply; NEMA 5-5P
OPVI	3269d	2	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	33"w-doors open 36.5" x 33 3/8"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	1	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.75"h x24"w x 19.43"d	E 120V: 13.9 amp
OPCI	4132	1	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	38.1"w x 22.8"h x 4"d	E 115V; cabling; Data

**ARCH RMNO 02-17-08**

**RMTITLE PUMP ROOM (AIR & VAC SYSTEM TO USE EXISTING EQUIP)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVCI	9901	1	Compressor, Dental Air: with switch & intake kit: EXISTING TO TRANSFER:DETAILS BY BENCO	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVCI	9902	2	Vacuum Pump System-Dry Vac: w/remote control panel 2 switch: EXISTING TO TRANSFER:	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVI	9903	2	Separator, Almagam for 10 to 20 Users: By Benco: Midmark HG5 HV: EXISTING TO TX: DETAILS BY	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVI	9904	2	Reverse Osmosis Univeral Dental Water Supply; High Efficiency: EXISTING TO TX: DETAILS BY	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS

**ARCH RMNO 02-17-09**  
**RMTITLE STORAGE, CYLINDER TANK**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	2425	1	Rack, Cylinder tank, to hold 6 cylinders; SPECS & DETAILS BY BENCO; WILL REQUIRE WALL	wall mount	30"w x 30"d x 36"h	NONE

**ARCH RMNO 02-17-10**  
**RMTITLE OPERATORY, CLOSED**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27"; DENTAL CLOSED OPS; IF NOT CABINET	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron, XRay Protective, Adult Dental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron, XRay Protective, Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT- REQUIRES BACKING; See Mfg	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPVI	6004	1	Microscope, Dental Procedure; A Series; SEE MFG DETAILS; ceiling mount structure to support 500 lbs	ceiling	4.70"w x 3.25"h x 3.0"d + column length	PER BENCO DRWGS
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted; Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control, PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate, Adj. Ht., (Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj. Ht., (Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO**

**02-17-11**

**RMTITLE**

**OPERATORY, CLOSED (Microscopic Procedures RM)**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24" w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-12**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6007	1	3 O'Clock Upper Storage, 3 door unit Wall mounted; Open Bay With Owner Preferred Dispensers-ADEC	wall	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-13**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data

OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC..: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO**                    **02-17-14**  
**RMTITLE**                        **OPEN OPERATORY BAY**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC..: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS

OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO**            **02-17-15**  
**RMTITLE**                **OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC.: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO**            **02-17-16**  
**RMTITLE**                **OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com



OPCI	4133	1 TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1 Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1 12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1 Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC.: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1 Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1 Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1 Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1 Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1 Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1 Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-17**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC.: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS



OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-18**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC.: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-19**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-20**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS

OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-21**  
**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS

OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14
--------	------	---	---	-----------------	-----------------	-------------------------------

**ARCH RMNO 02-17-22**

**RMTITLE OPEN OPERATORY BAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE;	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6008b	1	9 O'Clock, Accessory Console, 3 door unit Wall mounted;	floor	86x33x16	E, Data, H/C water
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6048	1	Light, Dental Operating, Single Head, Wall Mount	ceiling	6x48x24	See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**ARCH RMNO 02-17-24**

**RMTITLE LAB, DENTAL (CONFIRM PREFERRED CASEWORK SOURCE)**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1338	3	Stool, Backrest, Adj.Ht with footring, casters optional, for standing height workstations	floor	18"w x17"dx 42"h	None
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x36"h	None

OPOI	2944	1 Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3040a	1 Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	6010b	1 Dental Cabinetry System: Lab Area: PER PREFERRED MFG DESIGN	floor	PER ADEC DESIGN DRWGS	SEE DESIGN DRWGS
OPOI	6050	2 Macro Cab with Dust Collector: BY BENCO	countertop	12x12x15	E 115V
OPVI	6052	1 Mixer, Combination,Power/Vacuum,w/bench stand, complete: SEE BENCO	countertop	8x8x12	E 115V
OPVI	6054	1 Bin,Plaster,3 Compartment, 75# capacity (REQUIRES SUPPORT BACKING) SEE BENCO DETAILS	wall	21x8x18	None
OPSVI	6058	2 Pot, Curing, Pressure: 8 quart: By BENCO	countertop	12" diameter x 10"h	E 115V
OPSVI	6059	2 Knife, Proform Electric for dental labs: By BENCO	countertop	8"w x 6"d x 3"h	E 110V

**ARCH RMNO 02-17-25**  
**RMTITLE OFFICE, ASSISTANT SUPERVISOR**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO**            **02-17-26**  
**RMTITLE**            **OPEN WORK AREA, TEAM SPACE**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	8	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	36x12x65	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	TBD	None
OPVI	1500b	8	Workstation,Open Office Furniture System,Modular, Small Scale	floor	Various	E 115V Quad: Data-Com
OPOI	2941	8	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	8	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPVI	3062a	1	Wastebasket, Office,Tall, 23 gal, with flip lid (Blue Recycling-HIPPA)	floor	20lx11wx30h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPVI	3765a	1	Shredder, Box; with locking door & rigid liner; color/finish to be selected	floor	21.5"wx16"dx36"h	None
OPOI	3903	8	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3929	1	Copier, Multifunction, 44 ppm, Color/B/W (stand optional)	countertop	22.3"w x 23.4"d x 24.2"h	E 115V; Data/Com

**ARCH RMNO**            **02-17-28**  
**RMTITLE**            **OFFICE, DENTAL DIRECTOR**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	4	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1412	1	Table, Round, Diameter, 36"	floor	36 diamx30	None

OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO 02-17-29**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36x18x65	None
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO 02-17-30**

**RMTITLE EVS CLOSET**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2609	1	Rack, Wall Mounted to hold mops, etc...	floor	36"w x 1.5"d x 5"h	None
OPVI	3046	1	Trash Collector, 32 Gal Brute, On dolly, Round with wrap around caddy for janitorial services	floor	TBD	None

**ARCH RMNO 02-17-31**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OOOO	9999a	0	No Moveable Equipment Required	n/a	n/a	n/a

**ARCH RMNO 02-17-34**

**RMTITLE STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	9	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None

OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
------	------	---	---	-------	-------------	------

**ARCH RMNO 02-17-35**  
**RMTITLE OPERATORY, CLOSED**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14



**ARCH RMNO 02-17-37****RMTITLE PANORAMIC XRAY**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	1417	1	Table/Stand, For Computer, Printer, or opt paper storage below, On Casters	floor	26x24x30	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5761a	1	Radiographic, Dental, Panoramic-Digital:REQUIRES BACKING: EXISTING EQUIP TO TX	floor	See MFG Details	SEE PLAN MECA DETAILS
OPVI	5768a	1	Rack, Apron; Single; BACKING REQUIRED: BY BENCO	wall-Backing	3"H • 5"L prongs; on 16" centers	PER BENCO DRWG
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

**ARCH RMNO 02-17-38****RMTITLE STORAGE**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wtds,depths	floor	Various	None
OPVI	2254a	1	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None

**ARCH RMNO 02-17-39****RMTITLE OFFICE, MANAGER**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	24.5x28x37	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	24x24x37	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	36x18x64	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
CPCI	3115	1	Robe Hook, Double	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

**ARCH RMNO**            **02-17-40**  
**RMTITLE**                **LOCKERS AREA**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPVI	9997	1	Bench, Sitting Suitable for Location, Padded Seat, ADA/Bariatric	floor	Per Space Allowance	None

**ARCH RMNO**            **02-17-41**  
**RMTITLE**                **PANORAMIC XRAY**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	1417	1	Table/Stand, For Computer, Printer, or opt paper storage below, On Casters	floor	26x24x30	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14lx11wx16h	None
OPOI	3087	1	Dispenser, Glove for Single Box	wall mount	15"w x 5"d x 6"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5761	1	Radiographic, Dental, Panoramic-Digital: WALL MOUNT REQUIRES BACKING SEE PLAN MECA	floor	See MFG Details	SEE PLAN MECA DETAILS
OPVI	5768a	1	Rack, Apron; Single; BACKING REQUIRED: BY BENCO	wall-Backing	3"H • 5"L prongs; on 16" centers	PER BENCO DRWG
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

**ARCH RMNO**            **02-17-42**  
**RMTITLE**                **CLEAN ROOM-ICC**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19"w x 16"d x 32"h	None

OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPSVI	3100	2	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	2	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	5274	1	Incubator, Dry Heat, Open, 10/20 well, digital	countertop	11.3"w x 7.9"d x 3.9"h	E 115V
OPVI	6002	18	Amalgamator High Speed: BY BENCO	countertop	TBD	E 115V
OPSVI	6010a	1	Dental Cabinetry System: Clean Sterile: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPOI	6037	60	Handpiece, High Speed, Fiberoptic 5 hole: Impact Air 45:	stored	stored	None
OPSVI	6043	12	Intra Oral Camera-CS 1500: BY BENCO	attached	See Benco Details	None
OPSVI	6067	6	Sensor Kit Combo: MFG TBD; BENCO ITEM	stored	varies	None
OPSVI	6069	16	Sensor Kit Size 2: MFG TBD; BENCO ITEM	stored	varies	None
OPVI	6124	2	Cleaner/Lubricator, Handpiece, Air Driven, w/ handpiece: SEE BENCO DETAILS	countertop	7.5"w x 16.3"d x8.5"h	Air Supply 65 psi not to exceed 140 psi
OPVI	9221	2	Sterilizer, Table Top,Electric,Steam Type,w Printer & Pouch Rack: VOLTAGE TBD; SEE MFG DETAILS	countertop	23.8"w x17.8"d x17.8"h	E 115 OR 230 VAC, 50/60 Hz, 15 A Single Phase

**ARCH RMNO**            **02-17-43**  
**RMTITLE**             **SOILED ROOM-ICC**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19"w x 16"d x 32"h	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19"w x 16"d x 32"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: VERIFY OWNER SOURCE	wall mount	16"w x 18"h x 4"d	None
OPSVI	3100	2	Dispenser, Paper Towel: TYPE TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	2	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer	wall	TBD	None

OPSVI	6010e	1	Dental Cabinetry System: Soiled Room Processing: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPVI	6021a	2	Disinfectant, Thermal, for Dental Instruments, Undercounter: SEE BENCO DETAILS	undercounter	24x24x33.5	SEE BENCO DESIGN DRWGS
OPSVCI	6056	1	Autoclave:Single or Double Door-Pass Thru: PER MFG DESIGN	floor	28."w x 60"d x 75"h	SEE MFG DETAILS
OPVI	6407	1	Cleaner, Ultrasonic; With Digital Timer (BENCO) Drain, Mounting Rim: By BENCO	Countertop	See Benco Details	See Benco Details

**ARCH RMNO**            **02-17-44**  
**RMTITLE**            **OPEN WORK AREA, TECH**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPSVI	5050	2	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	30"w x 80"h x 29.5"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	6010f	1	Dental Cabinetry System: Tech Workspace: Design by Midmark & Benco	floor	See Midmark-Benco Details	See Midmark-Benco Details

**ARCH RMNO**            **02-17-46**  
**RMTITLE**            **STORAGE**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	2254	3	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf; various hts,wdts,depths	floor	Various	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, with drawer glide system & bins	floor	42"w x24"d x72"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None

**ARCH RMNO**            **02-17-47**  
**RMTITLE**            **OPEN WORK AREA, TECH**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPVI	0181	4	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5x28x48	None
OPOI	2941	4	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid	floor	20lx11wx30h	None
OPOI	3903	4	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

OPVI	3929	1	Copier, Multifunction, 44 ppm, Color/B/W (stand optional)	countertop	22.3"w x 23.4"d x 24.2"h	E 115V; Data/Com
OPSVI	6010f	1	Dental Cabinetry System: Tech Workspace: Design by Midmark & Benco	floor	See Midmark-Benco Details	See Midmark-Benco Details

**ARCH RMNO 02-17-48**  
**RMTITLE OPERATORY, CLOSED**

GROUP	EQNO	QTY	DESCRIPTION	LOC	SIZE	UTILITIES
OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27"; DENTAL CLOSED OPS; IF NOT CABINET	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron, XRay Protective, Adult Dental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron, XRay Protective, Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT- REQUIRES BACKING; See Mfg	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted; Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control, PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate, Adj. Ht., (Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj. Ht., (Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand 21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System; accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2; IEC/EN 60320-1 C14

**ARCH RMNO**            **02-17-49**  
**RMTITLE**            **OPERATORY, CLOSED**

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

**DEPTNO** 18  
**DEPTTITLE** CIRCULATION-2ND FLOOR  
**ARCH RMNO** 02-18-00  
**RMTITLE** CIRCULATION, 2ND FLOOR (AED LOCATION)

<b>GROUP</b>	<b>EQNO</b>	<b>QTY</b>	<b>DESCRIPTION</b>	<b>LOC</b>	<b>SIZE</b>	<b>UTILITIES</b>
OPOI	8202	1	Defibrillator, External, Automated (AED) Complete with Recessed Cabinet	Wall mount	15"w x 7"d x12"	None