

Equipment Planning

Cherokee Nation Health Center WPM STILWELL EXPANSION PROJECT

Prepared for Childers Architecture
May 22, 2020

Prepared by:
Interior Logistics



CPCI = CONTRACTOR PROVIDED-CONTRACTOR INSTALLED

OPCI = OWNER PROCURED-CONTRACTOR INSTALLED

OPVI = OWNER PROCURED-VENDOR INSTALLED

OPOI = OWNER PROCURED-OWNER INSTALLED

OPVICI = OWNER PROCURED-VENDOR & CONTRACTOR INSTALLED

OPSVI = OWNER PROCURED-SPECIALTY VENDOR INSTALLED

OPSVCI = OWNER PROCURED-SPECIALTY VENDOR & CONTRACTOR INSTALLED

Equipment Planning-Stilwell Expansion Project

DEPARTMENT TITLE: PUBLIC FACILITIES**ROOM NO 01-00-00****RMTITLE VESTIBULE, MAIN ENTRY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3044	1	Wastebasket, Decorative Type, Perforated Steel Panels Square Open Top, 20 Gallon Rigid Liner	floor	16.30"w x 16.30"d x27"h	None
OPSVI	3083	1	Station, Sanitizing Suitable for Gym Areas;Wipe & Hand Sanitizer Dispenser, Waste Container; Tex Blk	floor	47" H x 12" W x 17" D	None
OPVI	9996	2	Bench, Sitting Suitable for Location	floor	60"w x approx 20"d x 18"h	None

ROOM NO 01-00-01**RMTITLE LOBBY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1024	2	Seating, 3 Person, Couch Type	floor	74"w x29"d x30.75"h	None
OPVI	1026	2	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1052	1	Clock, Decorative, Battery Operated, Size TBD By Design	wall	16" Diameter	Battery Op
OPVI	1303	1	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	36" w x4"d x24"h	None: Consider Backing
OPVI	1410a	4	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	3044	1	Wastebasket, Decorative Type, Perforated Steel Panels Square Open Top, 20 Gallon Rigid Liner	floor	16.30"w x 16.30"d x27"h	None
OPCI	3909	1	Kiosk Center, Check In: Informational (???FUTURE PURCHASE)	floor	24"w x 42"h x 24"d	E 115V; Data Com
OPOI	8202	1	Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1	Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated

ROOM NO 01-00-02**RMTITLE REGISTRATION/RECEPTION 1ST FLOOR**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	0150	3 Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	3 Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPOI	2941	3 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPVI	3760	1 Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3903	3 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3936b	3 Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	3 Keyboard Tray	under desk	22" work track	None

ROOM NO 01-00-03
RMTITLE RESTROOM, WOMENS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-00-04
RMTITLE RESTROOM, MENS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-00-06
RMTITLE WAITING, PUBLIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1007	5	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None

OPVI	1019	11 Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1024	4 Seating, 3 Person, Couch Type	floor	74"w x29"d x30.75"h	None
OPVI	1026	16 Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1303	1 Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	36" w x4"d x24"h	None: Consider Backing
OPVI	1410a	11 Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	1421	2 Table, Sofa Suitable for Lounge/Waiting Areas	floor	30" diameter x 22"h	None
OPVI	3043	4 Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4136	2 Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

ROOM NO **01-00-07**
RMTITLE **LACTATION ROOM**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1005	1	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1754	1	Casework, Modular, Designed to Meet Space Req: LACTATION ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	2294	1	Diaper Pail, with Step	floor	12"w x 10"d x 32H	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
CPCI	3135	1	Diaper Changing Station, Horizontal Drop Down: TO BE PROVIDED & INSTALLED BY CONTRACTOR;	wall	35.25"L X3.75"D X22.25"H	None

ROOM NO **01-00-09**

RMTITLE **STORAGE**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO **01-00-11**

RMTITLE **CIRCULATION, CORRIDOR (BY PHARMACY)**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	8202	1	Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1	Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated

DEPARTMENT TITLE: MEDICAL RECORDS**ROOM NO** 01-01-00**RMTITLE** RECEPTION (2 STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0152	1	Chair, Rotary, Bariatric, Std.Back, On Casters, Adjustable Arms	floor	28"w x21"d x40.5-45.5"h	None
OPVI	0430d	2	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPOI	2941	2	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPVI	3760	1	Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3907	2	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936a	2	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	2	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-01-01**RMTITLE** OFFICE, FUTURE CHS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None

OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0431d	1	Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x20"d x26"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907	1	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936b	1	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

ROOM NO **01-01-02**
RMTITLE **OFFICE, CHS**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0431d	1	Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x20"d x26"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907x	1	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com

OPOI	3936b	1	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
------	-------	---	---	---------	------------------------	--------------

ROOM NO **01-01-03**

RMTITLE **OFFICE, CHS**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0431d	1	Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x20"d x26"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907x	1	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3913	1	Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm
OPOI	3936b	1	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

ROOM NO **01-01-04**

RMTITLE **OFFICE, PBC**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0151	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms, High Back	floor	27"w x 18"d x 40-45"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None

OPVI	0431d	1 Cabinet, Filing,Lateral, 2 drawer to fit under L desk top	floor	30"w x20"d x26"h	None
OPVI	1500	1 Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907x	1 Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936b	1 Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

ROOM NO 01-01-05

RMTITLE CIRCULATION: ALCOVE, COPY (AT END OF CORRIDOR)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm

ROOM NO 01-01-06

RMTITLE OFFICE, OPEN (X5 CLERICAL STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	5	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	5	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVICI	1500I	1	Workstation,Open Office Furniture System,Modular, 5 staff	floor	See Design	E Quad x 8; data/com x4
OPOI	2941	5	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907	5	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com

OPOI	3936a	5	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	5	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-01-07

RMTITLE CODER (X3 STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	3	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVICI	1500d	1	Workstation,Open Office Furniture System,Modular, 3 person	floor	Various	E Quad x 3: Data Com x 3
OPOI	2941x	3	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907x	3	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3936a	3	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	3	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-01-08

RMTITLE OFFICE, MEDICAL RECORDS STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0434d	1	Cabinet, Filing,Lateral, 4 drawer	floor	36"w x20"d x54"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-01-09

RMTITLE OFFICE, MEDICAL RECORDS STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0434d	1	Cabinet, Filing,Lateral, 4 drawer	floor	36"w x20"d x54"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-01-10

RMTITLE STORAGE, FILE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0435d	4	Cabinet, Filing, Lateral, 5 high-(optional flip top door for 5th space), Locking	floor	36"w x18.25"d x65.75"	None
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36'w x18'd x65'h	None
OPVI	1413d	1	Table, Work, Rectangular, Static, Standing Height	floor	48"l x18"d x36"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

DEPARTMENT TITLE: EMPLOYEE FACILITIES

ROOM NO 01-02-01
RMTITLE LOUNGE, STAFF (EMPLOYEE AREA)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1005	8	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVI	1411	2	Table, Square; With Conductor X Base	floor	36" Square x 30"h	None
OPVI	3062	2	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp;Water Supply;NEMA 5-15P
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	32.6" x 35"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details forDrain
OPVI	3349	2	Microwave Oven,Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.5"h x24"w x 18.5"d	E 120V: 13.9 amp

ROOM NO 01-02-02
RMTITLE RESTROOM, MENS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-02-03
RMTITLE RESTROOM, WOMENS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-02-04
RMTITLE TOILET, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None
------	-------	---	---	-------	---------------------	------

ROOM NO **01-02-05**

RMTITLE **TOILET, STAFF**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None
------	-------	---	---	-------	---------------------	------

ROOM NO **01-02-06**

RMTITLE **TOILET, STAFF**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None
------	-------	---	---	-------	---------------------	------

ROOM NO **01-02-07**

RMTITLE **TOILET, STAFF**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None
------	-------	---	---	-------	---------------------	------

ROOM NO **01-02-08**

RMTITLE **LOUNGE, STAFF**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	1005	12	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
------	------	----	---	-------	------------------	------

OPVI	1411	3	Table, Square; With Conductor X Base	floor	36" Square x 30"h	None
------	------	---	--------------------------------------	-------	-------------------	------

OPVI	3062	2	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
------	------	---	---	-------	------------------	------

OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
-------	------	---	---	------------	-----	------

OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
-------	------	---	--	------------	------------------	------

OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5" w x 17.75"d x 17"h	E 120V Dedicated; 15 amp;Water Supply;NEMA 5-15P
--------	------	---	---	------------	--------------------------	--

OPVI	3269d	2	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	32.6" x 35"d x 70"h	E 115V; Water Supply for Icemaker
------	-------	---	--	-------	---------------------	-----------------------------------

OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	2	Microwave Oven, Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.5"h x24"w x 18.5"d	E 120V: 13.9 amp

ROOM NO 02-19-07

RMTITLE WELLNESS CENTER, STAFF (FUTURE PURCHASE)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3062	1	Wastebasket, Office, Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3083	1	Station, Sanitizing Suitable for Gym Areas; Wipe & Hand Sanitizer Dispenser, Waste Container; Tex Blk	floor	47" H x 12" W x 17" D	None
OPVI	7103	1	Bike, Exercise, Upright, With Monitor: Precor (unit 169 lbs)	floor	48"l x21"w x62"h	E Ded 120V; 15 amp; Data
OPVI	7103c	1	Bike, Exercise, Recumbent; No step thru With Monitor: Precor (unit 219 lbs)	floor	67"l x23"w x54"h	E Ded 120V; 15 amp; Data
CPCI	7110c	3	Mirror(s), Wall (Size, Design Determined by Architect)	wall	TBD: SEE ARCHITECTURAL	None
OPVI	7128b	2	Treadmill, Individual Walking Exercise, W/Handrails	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data-Com
OPVI	7129a	1	Exercise Unit, Elliptical Trainer, w/progress monitor	floor	80"l x 28"w x 72h	E 120V Data-Com
OPVI	9996a	2	Bench, With 3 Cubbies, Suitable for Shoe Changing Areas in Gym	floor	36" Long x 13-1/2" Deep x 20" High	None

DEPARTMENT TITLE: FACILITIES MAINTENANCE

ROOM NO 01-03-01

RMTITLE FACILITIES MAINTENANCE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0212	80	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	22"w x20"d x32"h	None
OPVI	2201	1	Cabinet, Storage: Flammable Liquid: up to 30 gal: self closing door	floor	43"w x 18"d x 44"h	None
OPVI	2256c	20	Shelving Unit, Open Wire Shelves,Static; Starter Units Total of 4800 lb cap: Add on units 3000lb cap	floor	60"w x30"d x86"h	None
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x48"h	None
OPVI	3047	1	Trash Collector, 1/2 Cubic Yd, poly construction,color TBD on wheels with lid, color to be selected	floor	60"L X28"W X38.75"H	None
OPVI	3047a	2	Trash Collector w/Lid,and Dolly, 44 gallons (color TBD by Function)	floor	24"diameter x 37.5"h	None

DEPARTMENT TITLE: ENVIRONMENTAL SERVICES: HOUSEKEEPING

ROOM NO 01-04-01

RMTITLE BIO-HAZARD

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3032	1	Container, Bio Hazard; Rated-Med Waste: Hinged Lid; Red 65 Gal Cap; mobile	floor	32.3"l x 25.3"w x 41.8"h	None

ROOM NO 01-04-02

RMTITLE EVS CLOSET

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2256b	1	Shelving Unit, Open Wire Shelves, 4 Shelf Unit, Static	floor	48"w X24"d x74"h	None
OPVI	3046	1	Trash Collector, 32 Gal Brute, Round, with lid, dolly & wrap around caddy for janitorial services	floor	22"diameter x37.5"h with dolly & lid	None
OPVI	3047a	1	Trash Collector w/Lid,and Dolly, 44 gallons (color TBD by Function)	floor	24"diameter x 37.5"h	None
OPVI	3079	1	Bucket, Mop with ringer (SIZE & COLOR TBD)	floor	TBD	None

ROOM NO 01-04-03

RMTITLE EVS BULK STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2256b	2	Shelving Unit, Open Wire Shelves, 4 Shelf Unit, Static	floor	48"w X24"d x74"h	None
OPVI	3046	2	Trash Collector, 32 Gal Brute, Round, with lid, dolly & wrap around caddy for janitorial services	floor	22"diameter x37.5"h with dolly & lid	None
OPVI	3079	1	Bucket, Mop with ringer (SIZE & COLOR TBD)	floor	TBD	None

ROOM NO 01-04-04

RMTITLE OFFICE, EVS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	0150	1 Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	1 Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1 Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1 Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3913	1 Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm

ROOM NO 01-04-05
RMTITLE LINEN, CLEAN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3005	2	Cart, Linen, w/ 4 shelves and Staph Check enclosure	floor	50"w x 29"d x 69"h	None

ROOM NO 01-04-06
RMTITLE DRYER, LAUNDRY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1413d	1	Table, Work, Rectangular, Static, Standing Height	floor	48"l x18"d x36"h	None
OPCI	3021	1	Dryer, Front Load, Commercial, Electric, 7 Cu Ft Capacity	floor	26.88"w x28" install depth x44.34"h	E 240/60/1, 3-prong cord, 30 AMP; SEE ADD DETAILS
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO **01-04-07**
RMTITLE **WASHER, LAUNDRY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3017	1	Hamper, Laundry: round, mobile	floor	24"diameter x 34"h	None
OPCI	3020	1	Washer, Extractor, Front Load, Commercial, Energy Star Rated, 20 lb capacity	floor	27"w x27.6"d x42.5"h	E 208-240/60/1, 2W+G,20AMP; SEE ADDITIONAL DETAILS
OPVI	3023	1	Dispensing System, Laundry Detergent (LIKELY TO BE SOURCED THRU OWNER JANITORIAL SUPPLIER)	wall mount	19.75"l x 6.75"w x29"h	E 115V?; Water Supply
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None

DEPARTMENT TITLE: MEP SPACE**ROOM NO** 01-05-01**RMTITLE** MEP LARGE SPACE, 1ST FLOOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3045	1	Wastebasket, 25 gallon, with rigid Liner, UL rated,	floor	18" diameter x 42"h	None
OPVI	3972	1	Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly	floor	38"l x 19/25"w x33.5"h	None

ROOM NO 01-05-03**RMTITLE** I.T. ROOM, 1ST FLOOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2603	1	Bench, Work for I.T. Technician Repair Area: Lights/Bins/Power/Data Strip	floor	60"w x 24"d x Adj Ht	E 115V x 2 Quad: 2 Data-com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3972	1	Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly	floor	38"l x 19/25"w x33.5"h	None

ROOM NO 01-05-04**RMTITLE** I.T. ROOM, 1ST FLOOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3972	1	Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly	floor	38"l x 19/25"w x33.5"h	None

ROOM NO 02-05-10**RMTITLE** I.T. ROOM, 2ND FLOOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3972	1	Cart, Trades Utility; for Facilities Maintenance or I.T. Dept, w/doors/drawers/poly	floor	38"l x 19/25"w x33.5"h	None

DEPARTMENT TITLE: WINGS

ROOM NO **01-06-01**
RMTITLE **OFFICE, WINGS**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18"d x65"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1005	1	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVICI	1500c	1	Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	2	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

ROOM NO **01-06-02**
RMTITLE **OFFICE, WINGS**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None

OPVI	0680	1 Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36'w x18'd x65'h	None
OPVI	0700	1 Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1005	1 Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVICI	1500c	1 Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	2 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

DEPARTMENT TITLE: PHYSICAL THERAPY**ROOM NO** 01-07-00**RMTITLE** WAITING, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1007	3	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVI	1019	6	Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1303	1	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	36" w x4"d x24"h	None: Consider Backing
OPVI	1410a	1	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None

ROOM NO 01-07-01**RMTITLE** RECEPTION, PT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

OPOI 3916 1 Printer, Multi Function, Fax/Copy/Scan countertop 24.5"w x 19"d x 25.5"h E 115V:Data Com

ROOM NO 01-07-03

RMTITLE OFFICE, PT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913	1	Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm

ROOM NO 01-07-04

RMTITLE WORKSPACE, PT ASSISTANT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1500a	1	Workstation, Furniture System, With Transaction Top: Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPVI	2256a	1	Shelving Unit, Mobile, Adj Shelves, Bins, Bumper, Door Type TBD	floor	37.5"wx28.75"d x75.25"h	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant, 28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-07-05

RMTITLE PATIENT TOILET & DRESSING ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
CPCI	1805	1	Mirror, Full Length	wall	18x.5x60	None
OPVI	3040a	1	Wastebasket, Step-On, Plastic, Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPVI	9996	1	Bench, Sitting Suitable for Location	floor	60"w x approx 20"d x 18"h	None

ROOM NO 01-07-06

RMTITLE PATIENT TOILET & DRESSING ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
CPCI	1805	1	Mirror, Full Length	wall	18x.5x60	None
OPVI	3040a	1	Wastebasket, Step-On, Plastic, Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	

OPVI	9996	1	Bench, Sitting Suitable for Location	floor	60"w x approx 20"d x 18"h	None
------	------	---	--------------------------------------	-------	---------------------------	------

ROOM NO 01-07-07

RMTITLE TREATMENT, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	7115b	1	Table, Treatment; EXISTING EQUIPMENT TO TX	floor	TBD	E 115V
OPVI	7118	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart: EXISTING TO TX	floor	20"w x15"d x48"h	E 115V

ROOM NO 01-07-08

RMTITLE TREATMENT, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	Overall Width 25": Seat 16" Diameter	None

OPVI	1750	1 Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	7115b	1 Table, Treatment; EXISTING EQUIPMENT TO TX	floor	TBD	E 115V
OPVI	7118	1 Stimulator, Multi and Ultrasound, 2 Channel on portable cart: EXISTING TO TX	floor	20"w x15"d x48"h	E 115V

ROOM NO 01-07-09
RMTITLE TREATMENT, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1333	1	Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	7115a	1	Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7118a	1	Stimulator, Multi and Ultrasound, 2 Channel on portable cart	floor	17"w x16.25"d x44"h	E 115V

ROOM NO 01-07-10
RMTITLE STORAGE, PT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3005a	1	Cart, Linen, w/ 4 shelves and Staph Check enclosure	floor	36"w x 18"d x 69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	7108	1	Chilling Unit, Cold Packs, Complete w/packs, side table	floor	27"l x 16"d x 31"h	E 115; near drain
OPVI	7108a	1	Hydrocolator, Hot Pack, Unit: EXISTING EQUIPMENT TO TX	floor	27"l x 16"d x 33"h	E 115; near drain
OPVI	7108b	2	Rack for Hydrocolator towels/covers, folding hooks, 6/rack: REQUIRES BACKING	wall	36x15x2	None
OPCI	7122	1	Rack, Wall for Cane & Crutch, capacity of 6 crutch sets per support: REQUIRES BACKING	wall/backin g	66"l x5.25"w x with hooks 9"d	None:Requires Backing
OPOI	9858a	1	Wheelchair: EXISTING EQUIPMENT TO TX	floor	24x30x42	None

ROOM NO 01-07-11
RMTITLE TREATMENT PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None

OPVI	1333	1 Stool, Clinical, with Backrest, Pneumatical Adjustable Base	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1 Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	7115a	1 Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7118a	1 Stimulator, Multi and Ultrasound, 2 Channel on portable cart	floor	17"w x16.25"d x44"h	E 115V

ROOM NO **01-07-12**
RMTITLE **STORAGE, SOILED LINEN**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3016	2	Hamper, Linen, Mobile, Foot Operated Lid, Space Saver, Fits in Corners	floor	18"w x 18"l x 32"h	None

ROOM NO **01-07-13**
RMTITLE **NOURISHMENT AREA**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None

OPSVCI	3307	1 Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details forDrain
OPSVI	3310	1 Refrigerator, Undercounter, ADA 5.5 cubit ft, auto defrost, SS	floor	23.63"w x 23.63"d x 32"h	E 115V
OPCI	4136	1 Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data
OPCI	4138	1 Wall Mount for TV Monitors up to 55"; movement action TBD: OWNER TO SPECIFY: REQUIRES BACKING	wall	See Details	None
OPOI	4144	1 Home Video Game Console (Wii station): EXISTING EQUIPMENT TBD?????	countertop	3"w x 9"d x 8"h	video cable connection

ROOM NO **01-07-14**
RMTITLE **EXERCISE AREA**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1006	5	Chair, Plastic, 4 legs, stackable	floor	approx 20'w x21'd x32h	None
OPVI	1335a	4	Stool, No Backrest, Adj. Ht. Casters, Foot ring (22"-32" adj ht)	floor	Overall Width 22": Seat 16"Diameter	None
OPVI	2521	1	Board, Combination, Bulletin-Chalkless Marker 50/50	wall	36"w x1"d x24"h	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	3	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPSVI	3083	1	Station, Sanitizing Suitable for Gym Areas;Wipe & Hand Sanitizer Dispenser, Waste Container; Tex Blk	floor	47" H x 12" W x 17" D	None
OPSVI	3102a	3	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPCI	4136	1	Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data
OPCI	4136a	2	Television, Flat Panel, 55" Ceiling Mount: OWNER TO SPECIFY: REQUIRES ABOVE CEILING SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data
OPCI	4138	1	Wall Mount for TV Monitors up to 55"; movement action TBD: OWNER TO SPECIFY: REQUIRES BACKING	wall	See Details	None

OPCI	4138a	2 Ceiling Mount for TV Monitors up to 55"; movement action TBD: REQUIRES BACKING	wall	See Details	None
OPVI	7102a	1 Upper Body Exerciser-Step Thru Seating: SciFit Pro 1	floor	61"l x 30"w x 77"h	E Ded 115V; Data
OPVI	7103a	1 Bike, Exercise, Recumbent, Walk Thru Design, With Monitor; (For PT Area)	floor	64"l x 25.4"w x 47.5"h	E Ded 120V; 15 amp; Data
OPVI	7103c	1 Bike, Exercise, Recumbent; No step thru With Monitor: Precor (unit 219 lbs)	floor	67"l x23"w x54"h	E Ded 120V; 15 amp; Data
OPVI	7103d	2 Bike, Air: Assault	floor	51"l x23.5"w x35.5"h	None
OPOI	7107	1 Exercise Unit, Upper and Lower Body, BowFlex XLT Gym: EXISTING EQUIPMENT TO TX	floor	98lx50wx83h	None Req
OPOI	7107b	1 Exercise Unit, Total Gym, Fit: CLINICAL PKG	floor	106"l x 38"w x65"h	None
OPVI	7110a	1 Mirror, Glass, Tempered, 3 Section, Mobile: EXISTING EQUIPMENT TO TX	floor	30"w x10"d x72"h	None
OPVI	7110b	1 Storage, Multipurpose, for weights, includes rack, peg hooks: EXISTING TO TX	floor	24x18x72	None
OPVI	7113	1 Parallel Bars, Powered, Adult Adj Ht. & Width, 10' long: EXISTING EQUIPMENT TO TX	floor	up to 25"w x 10'l x42"h	E 120V-5 Amp: floor outlet
OPVI	7115a	1 Table, Treatment, Tilt, Hi/Low, Powered, 3 section	floor	78"l x30"w x adj ht	E 115V
OPVI	7120a	1 Rack, Combination weight and dumbbells-opposite sides	floor	31"w x53"h x24"d	None
OPVI	7120b	1 Rack, Progression Weight, Static With 1 Pair 5#, 10#, 15#, 20#, 25# & 30# Dumbbells	floor	62"w x 23"d x 32"h	None
OPVI	7128b	1 Treadmill, Individual Walking Exercise, W/Handrails	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data-Com
OPVI	7128c	1 Treadmill, Individual Walking Exercise, W/Handrails: EXISTING EQUIPMENT TO TX	floor	80"l x39.5"w x73.5"h	E Dedicated 120V; Data-Com
OPVI	7129a	1 Exercise Unit, Elliptical Trainer, w/progress monitor	floor	80"lx 28"w x"72h	E 120V Data-Com
OPVI	7129b	1 Stepper, Recumbent: Sci Fit Step1; EXISTING EQUIPMENT TO TX	floor	68"l x30"w x48"h	E 115V; Data

OPVI	7135	1 Dynamometer, Hand, Hydraulic: EXISTING TO TX	stored	n/a	None
OPVI	7136	1 Biomechanical Ankle Platform System (BAPS): EXISTING TO TX	stored	n/a	None
OPVI	7137b	1 Rack for Storing Therapy Balls-Mobile	floor	64"l x 19"d x64"h	None
OPVI	7163	1 Body Solid Universal Gym	floor	12' w x 12'd x unknown ht	None
OPOI	8202	1 Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1 Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated

DEPARTMENT TITLE: BEHAVIORAL HEALTH**ROOM NO 01-08-00****RMTITLE WAITING, PATIENT**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1010	2	Chair, Child	floor	15.25"w x15"d 24"h	None
OPVI	1025	2	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	4	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1409	1	Table, Child Use, Square or Round (optional shape & colors)	floor	24" square x22"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4132	1	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	37.8" w x 22.1" h x2.2" d	E 115V; cabling; Data

ROOM NO 01-08-01**RMTITLE RECEPTION, BEHAVIORAL HEALTH**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None

CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

ROOM NO 01-08-02

RMTITLE CIRCULATION (AED LOCATION) BETWEEN PC-EYE CLINICS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	8202	1	Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1	Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated

ROOM NO 01-08-03

RMTITLE OFFICE, THERAPIST

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-04

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-08-05

RMTITLE OBSERVATION, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	1	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0430d	2	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPVI	2810	1	Vacuum, Hand, Portable, Rechargeable	stored	16"l x 9"w x 10"h	E 115V for charging
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	

OPOI 3903 1 Computer,Personal & Monitor for Desk desktop 24"w x 18"d x 22"h E 115V; Data Com
Use: TO BE SPECIFIED BY OWNER

ROOM NO 01-08-06

RMTITLE PLAY THERAPY (Panic Button)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0681	3	Cabinet Suitable for display of teaching materials or models, locking doors optional	floor	36"w x 23"d x 65"h	None
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1010	4	Chair, Child	floor	15.25"w x15"d 24"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1092	1	Rug, Area: Easy Clean up: Childrens Theming	floor	60"w x78"l	None
OPVI	1409	1	Table, Child Use, Square or Round (optional shape & colors)	floor	24" square x22"h	None
OPVI	1419	1	Table, Rectangular, On Casters, Optional Flip Top, Optional Drawer	floor	30"w x 24"d x 30"h	None
OPVI	2268	3	Shelves, Suitable for storage of toys, games, must be child accessible, secure to wall	floor/secure	34"w x 13"d x 19"h	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None

ROOM NO 01-08-07

RMTITLE OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none

OPVI	1025	1 Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1 Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1 Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1 Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-08

RMTITLE OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op

OPVI	1500	1 Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-09

RMTITLE OFFICE, THERAPIST (Panic Button & DIMMABLE LIGHT)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None

OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-10

RMTITLE GROUP THERAPY ROOM (Panic Button- DIMMABLE LIGHT)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0212	12	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	22"w x20"d x32"h	None
OPVI	0552	1	Credenza, w/4 doors, 2 drawers	floor	72"w x 24"d x36"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1402a	4	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"l x 24"d x30"h	E 115V-floor: Data-Com
OPVI	2524	1	Board, Marker, Cleanable, OPTIONAL AS PORTABLE	wall	36"w x 72"h	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPOI	2945	1	Telephone, Conference: TO BE SPECIFIED BY OWNER	desktop	14X13X3	E 115V; Data-Com
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVI	3216d	1	Coffeemaker,Pour Over, 12 cup, always ready, 1 burner; with 2 EASY POUR DECANTERS	countertop	8"w x17.7"d x17.3"h	E Ded 120V
OPSVCI	3307a	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 10# ice bin, up to 315# ice in 24 hr period	countertop	16.25"w x24"d x35"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain

OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4136	1	Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

ROOM NO 01-08-11

RMTITLE OFFICE, MEDICAL SOCIAL WORKER (Panic Button)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary, Std. Back, 5-Legged, Ergonomic, Adj. Short Arms	floor	27"w x 18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x 20"d x 32"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant, 28qt	floor	14"l x 11"w x 16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-12

RMTITLE OFFICE, APRN (Panic Button)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary, Std. Back, 5-Legged, Ergonomic, Adj. Short Arms	floor	27"w x 18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x 20"d x 32"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-08-13

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18"d x65"h	None
OPVI	2254	4	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-08-14

RMTITLE OFFICE, THERAPIST, BEHAVIORAL HEALTH

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-08-15

RMTITLE OFFICE, THERAPIST, BEHAVIORAL HEALTH

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0383	1	Bookcase,Non-Sectional,Open, 3 Adj. Shelves	floor	30"w x15"d x66"h	None
OPVI	0701	1	Bookcase/Lateral file/Tower combination unit	floor	36"w x 24"d x 66"h	none
OPVI	1025	1	Seating, 2 Person, Couch Type	floor	52"w x29"d x30.75"h	None
OPVI	1026	1	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None

CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-08-16

RMTITLE EXAM, BEHAVIORAL HEALTH

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On, Plastic, Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On, Plastic, Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap: Type TBD BY OWNER- CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging

OPOI	3903	0 Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation: SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x 8"l x 10"h	None
OPVI	8572	1 Light, Examination, Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1 Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312a	1 Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer, recep	floor	28"w x 18"-37"h x 76" ext'd	E 115VAC, 50/60Hz, 8amps
OPVI	9854	1 Sundry Jar Set of 5, for Exam Room Supplies (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x 3.9w x 7"h	None
CPCI	9993	1 Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

DEPARTMENT TITLE: PRIMARY CARE**ROOM NO 01-09-00****RMTITLE RECEPTION, PRIMARY CARE: 5 STAFF**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	5	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj,Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	5	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPOI	2941	5	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	5	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907	5	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3916	2	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPOI	3936b	5	Scanner, Desktop, Auto document feeder, Duplex, Color	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	5	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-09-02**RMTITLE ALCOVE, WEIGHT**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1758	1	Casework, Modular, Designed to Meet Space Req: WEIGHT ALCOVE: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3041a	1	Wastebasket, End Step-On, Plastic, Rectangular, Color Beige, 8 Gallon	floor	approx 12.3"w x19.6"d x22.3"h	None
OPVI	8911	1	Scale,Person Weighing, 1000 lb capacity, for wheelchair access, with stadiometer	floor	28"w x32"dx 55"h	E 115V; or battery op

ROOM NO 01-09-03
RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-09-04
RMTITLE EXAM, PC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging

OPOI	3903	0	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPVI	8572	1	Light, Examination,Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1	Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312	1	Table, Patient Examination, Powered (Midmark 626) 32" width seat with built in I.Q. Scale System	floor	32"w x 60"-76"l x 15.5"-37"h	E 115VAC
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO **01-09-05**
RMTITLE **EXAM, PC**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None

OPVI	3041	1 Wastebasket, Step-On, Plastic, Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1 Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	0 Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPVI	8572	1 Light, Examination, Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1 Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312a	1 Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer, recep	floor	28"w x18"-37"h x76" ext'd	E 115VAC,50/60Hz, 8amps
OPVI	9854	1 Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None
CPCI	9993	1 Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO 01-09-06
RMTITLE EXAM, PC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	0	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com

OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPVI	8572	1	Light, Examination, Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1	Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312a	1	Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer, recep	floor	28"w x18"-37"h x76" ext'd	E 115VAC,50/60Hz, 8amps
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO 01-09-07

RMTITLE MEDICATION ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3062	1	Wastebasket, Office, Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap: Type TBD BY OWNER- CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

OPOI	5018	1 Sink, Smart for drug waste containment	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPSVI	5030b	1 Controlled Dispensing System - 2 Drawer Main-Countertop Unit (LEASE ITEM?)	Countertop	23"W x27"D x29"H	E 115V, Data-Com x 2
OPSVI	5050	1 Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	29"w x 79.6h"h x 34.2"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1 Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 32"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPVI	9299	1 Stand, IV, with 4 hangers, 5 legged, mobile	floor	16" diameter	None

ROOM NO **01-09-11**
RMTITLE **TREATMENT ROOM**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1335	1	Stool, Clinical, NO Backrest, Pneumatical Adjustable Base, Optional Footring	floor	Overall Width 22": Seat 16"Diameter	None
OPVI	1752	1	Casework, Modular, Designed to Meet Space Req: TYPICAL TREATMENT ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8166	1	Cart, Procedures Set Up (to be part of the casework package from Midmark)	floor	approx 24.25"w x 30"l x38.25"h	None
OPVI	8261	1	Electrosurgical Unit, (Hyfrecator) Minor Procedures - Cautery, on stand	floor	18"diam x60	E 115V
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPCI	8570b	1	Light, Minor Surgical,Ceiling Pedestal Mount Single Arm: CEILING SUPPPORT REQUIRED	ceiling	22 diam., 30-54	E,Struc.Support, 120VAC
OPSVI	8705	1	Monitor, Vital Signs; BP, Pulse, SPO2, Resp, EKG, On Optional Cart	stand or cart	10.2"w x8.6"h x5.7"d	E 115V; Data/Com
OPVI	8791	1	Oximeter, Pulse, Handheld/Portable	stored	1.5"w x 3.25"l	battery operated
OPVI	8801	1	Pump, Infusion on Mobile Stand; MFG & MODEL TBD	floor	20diam x65	E 115V(recharging)
OPVI	9301	1	Stand, Mayo, Removable Tray,Adj.Ht.	floor	16.25"w x 21.25"l x 37-61" h	None
OPVI	9312c	1	Table, Universal Procedure,(lithotomy/trendelenberg/podiatry)	floor	28"-31"w x18"-40"h x72	E,115VAC, 10amp, 50/60Hz
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO 01-09-12

RMTITLE CLEAN ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1757	1	Casework, Modular, Designed to Meet Space Req: CLEAN ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	5274	1	Incubator, Dry Heat, Open, 10/20 well, digital	countertop	11.3"w x 7.9"d x 3.9"h	E 115V
OPVI	9221	1	Sterilizer, Table Top,Electric,Steam Type,w Printer & Pouch Rack: SEE MFG DETAILS	countertop	23.8"w x17.8"d x17.8"h	E Ded 115 VAC, 50/60 Hz, 15 A Single Phase

ROOM NO **01-09-13**
RMTITLE **SOILED ROOM**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1756	1	Casework, Modular, Designed to Meet Space Req: SOILED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3016	1	Hamper, Linen, Mobile, Foot Operated Lid, Space Saver, Fits in Corners	floor	18"w x 18"l x 32"h	None
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

ROOM NO **01-09-14**
RMTITLE **STORAGE, CLEAN**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	6	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-09-15
RMTITLE OFFICE, FLOATING FOR 8 PROVIDERS

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	8	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	8	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1303	1	Rack, Pamphlet/Magazine w/adj dividers: BACKING REQUIRED	wall	36" w x4"d x24"h	None: Consider Backing
OPVICI	1500g	1	Workstation,Open Office Furniture System,Modular, 8 staff	floor	See Design	E Quad x 8; data/com x4
OPOI	2941	8	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	8	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903	8	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	8	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-09-16
RMTITLE OFFICE, PC PHYSICIAN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-09-17

RMTITLE OFFICE, PC PHYSICIAN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-09-18

RMTITLE OPEN WORK, TEAM SPACE: 6 STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	6	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	6	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1406	1	Common Work Surface/Table Top Suitable for Function of Room/will have storage underneath	floor	72"lx36"d x42"h	None

OPVI	1500	6 Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	6 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	6 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903	6 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	6 Keyboard Tray	under desk	22" work track	None

DEPARTMENT TITLE: EYE CARE**ROOM NO 01-10-00****RMTITLE DISPENSING, EYEWEAR**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPVI	3905	2	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPSVI	6400	1	Eye Wear Dispensing Furnishing Package by Fashion Optical: SEE FASHION OPTICAL DETAILS	floor	See Design	See Design
OPVI	6402	2	Chair, Optometric Dispensing Technician, w/Back,Casters	floor	24x24x30	None
OPVI	6402a	4	Chair, Optometric Dispensing, No arms, For Patient Use	floor	24x24x30	None

ROOM NO 01-10-00a**RMTITLE WORK AREA, TECHNICIAN**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0182	1	Chair, Rotary,Std.Back,4-Legs, Nesting,Ergonomic, Partial Arm, Standing Height	floor	24.5"w x28"d x48"h	None
OPVI	1202	2	Lighting, Task, Overhead, LED, Warm Light (electrical behind upper cabinets)	under overheads	31"W	E 115V Receptacle Behind Uppers
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

OPOI	6412	1	Safe, Small, for storing moneys; EXISTING EQUIPMENT TO TX		TBD	None
OPVI	6450	1	Warming Unit,for Frame Adjustment, Hot Air Warmed	desktop	8"w x5.25"l x8.5"h	E 110V
ROOM NO		01-10-02				
RMTITLE		TESTING, PRELIMINARY				
GROUP	EQNO	QTY	DESCRIPTION	LOCATION	SIZE	UTILITIES
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	6411	1	Cart, Supply-Portable, Soft Edge w/Handle Top;Optional lock;(TO BE PART OF THE MIDMARK CASEWORK	floor	18"w x 18"d x33"h	None
OPVI	6420	1	Ultrasound, B Scan, handheld	stored	n/a	Battery Operated
OPVI	6421	1	Pachymeter, System, blue tooth technology, printer (EXISTING EQUIPMENT TO TX)	countertop	printer 4"w x 6"d x1.54"h	E 115 for printer & charging
OPVI	6422	1	Refractor/Keratometer/Topographer, Automated; Sits on a table top #6439	tabletop	13.5"w x 21.18"d x18.77"h	E-Ded: Data
OPVI	6437	1	Stool, Backrest, Adj.Ht., On Casters (Is foot ring required)	floor	15"diameter x 19-27" ht	None
OPVI	6439	1	Tables, Instrument,Casters, Motorized, height adjustable,	floor; mobile	36"wx 24"d x adj ht	E-Dedicated, 1 Data
OPVI	6444b	1	Visual Field Analyzer, Automated Humphrey Matrix 800: EXISTING EQUIPMENT TO TX	shared table to	12"w x 22"d x17"h	E-115V Dedicated; Data
OPVI	6446	2	Tonometer, Hand Held	stored	n/a	None

OPVI	6489	0	Lensometer, Computerized	countertop	6.3"w x 16.53"d x12.2"h	E 115; Data
OPVI	6490	1	Lensometer, Computerized w/printer; TO DETERMINE IF TABLE COMES WITH UNIT	tabletop	48"w x 24"d x adj ht	E 115; data com
OPVI	6501	1	Monitor, Blood Pressure on stand (EXISTING EQUIPMENT TO TX)	floor	18" diameter	E 115V for charging
OPVI	6509	1	Tester, Brightness Acuity	stored	n/a	None
OPVI	6510	1	Tests, Optometric, Various: EXISTING EQUIPMENT TO TRANSFER	stored	n/a	None
OPVI	6528	1	Test, Color Vision: HHR standard version/with illuminator (EXISTING EQUIPMENT TO TX)	stored	6"w x 5"h x .5"d	None

ROOM NO 01-10-03

RMTITLE EXAM, EYE (RM W/CHAIR MOVER)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

OPCI	4133a	1 Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1 Chair, Eye Exam, Powered	floor	22.05"w x48" in upright position xadj ht	E Ded 110V, Lighting Control
OPVI	6406a	1 Floor Unit, Stand & Instrument Control	floor	21.25"w x 21.25"d x74.44"h	E-115V
OPVI	6406c	1 Chair Glide, Mover, To Accommodate Wheelchairs, for bases up to 22.5" wide; under exam	under chair	52"l x 26"w x 0.75"h	None
OPVI	6417	3 Handles, Rechargeable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1 Refractor for Stand	attached	n/a	None
OPVI	6425	1 Ophthalmoscope, Indirect, Binocular	stored	n/a	Battery Operated
OPVI	6425a	1 Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1 Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPOI	6429a	1 Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data
OPVI	6430	1 Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1 Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1 Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand
OPVI	6437	2 Stool, Backrest, Adj.Ht., On Casters (Is foot ring required)	floor	15"diameter x 19-27" ht	None
OPVI	6445	1 Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1 Tonometer Mount	attached	N/A	None
OPVI	6448	1 Frame, Trial Lens, Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None

OPVI	6449	1 Trial Lens Set, Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPOI	6452	1 Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software:Wall Mount: BACKING REQ	wall	TBD	E 115V: Data-Com
OPVI	6481	1 Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1 Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1 Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1 Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1 Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1 Forceps, Jeweler	stored	n/a	None
OPVI	9127	1 Sphygmomanometer, Aneroid, Hand Use, Portable, With Assortment	stored	n/a-stored item	None

ROOM NO 01-10-04

RMTITLE EXAM, EYE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step-On, Plastic, Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None

OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPCI	4133a	1 Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1 Chair, Eye Exam, Powered	floor	22.05"w x48" in upright position xadj ht	E Ded 110V, Lighting Control
OPVI	6406a	1 Floor Unit, Stand & Instrument Control	floor	21.25"w x 21.25"d x74.44"h	E-115V
OPVI	6417	3 Handles, Rechargeable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1 Refractor for Stand	attached	n/a	None
OPVI	6425	1 Ophthalmoscope,Indirect,Binocular	stored	n/a	Battery Operated
OPVI	6425a	1 Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1 Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPOI	6429a	1 Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data
OPVI	6430	1 Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1 Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1 Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand

OPVI	6437	2 Stool, Backrest, Adj.Ht., On Casters (Is foot ring required)	floor	15"diameter x 19-27" ht	None
OPVI	6445	1 Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1 Tonometer Mount	attached	N/A	None
OPVI	6448	1 Frame,Trial Lens,Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None
OPVI	6449	1 Trial Lens Set,Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPOI	6452	1 Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software:Wall Mount: BACKING REQ	wall	TBD	E 115V: Data-Com
OPVI	6481	1 Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1 Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1 Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1 Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1 Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1 Forceps, Jeweler	stored	n/a	None
OPVI	9127	1 Sphygmomanometer,Aneroid,Hand Use,Portable, With Assortment	stored	n/a-stored item	None

ROOM NO 01-10-05

RMTITLE SPECIAL TESTING

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None

OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	6403	1	Ocular Coherence Tomography, with monitor, software & table (EXISTING EQUIPMENT TO TX)	tabletop	table: 39"l x 22"w x adj ht	E 115V-DED, Data-Com
OPVI	6404	1	Camera, Fundal/Pt Mgmt Software System/on table (EXISTING EQUIPMENT TO TX)	sits on table	table: 36"w x 20"d x adj ht	E 115V,-Dedicated, Data-Com
OPVI	6437	2	Stool, Backrest, Adj.Ht., On Casters (ls foot ring required)	floor	15"diameter x 19-27" ht	None
OPVI	6444	1	Visual Field Analyzer, Automated; SITS ON ITS OWN TABLE: EXISTING EQUIPMENT TO TX	tabletop	36"w x24"d x ADJ HT	E 115V-Dedicated; Data

ROOM NO 01-10-06

RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

ROOM NO 01-10-07

RMTITLE EXAM, EYE (RM W/CHAIR MOVER)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None

OPOI	3087	1 Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPCI	4133a	1 Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPVI	6406	1 Chair, Eye Exam, Powered	floor	22.05"w x48" in upright position xadj ht	E Ded 110V, Lighting Control
OPVI	6406a	1 Floor Unit, Stand & Instrument Control	floor	21.25"w x 21.25"d x74.44"h	E-115V
OPVI	6406c	1 Chair Glide, Mover, To Accommodate Wheelchairs, for bases up to 22.5" wide; under exam	under chair	52"l x 26"w x 0.75"h	None
OPVI	6417	3 Handles, Rechargeable, Contained in wells on stand	stored	n/a	E 115V:(located at base)
OPVI	6424	1 Refractor for Stand	attached	n/a	None
OPVI	6425	1 Ophthalmoscope,Indirect,Binocular	stored	n/a	Battery Operated
OPVI	6425a	1 Power Source, BIO & wall rack for equipment #6425; wall hung	wall mount	8"w x 8"h x 5"d	E 115V above counter/wall mount
OPVI	6426	1 Ophthalmoscope, Direct Coax	stored in stand	n/a	Stand Charge
OPOI	6429a	1 Visual Acuity System, All In One , Wall Mount w/ Remote Control: BACKING REQUIRED	wall mount	18.6"w x15"h x 2.9"d	E 115V-Data

OPVI	6430	1 Transilluminator, 3.5V Halogen	stored in stand	n/a	None
OPVI	6431	1 Retinoscope, Streak, Welch Allyn Elite	stored in stand	varies	None
OPVI	6435	1 Lamp, Slit: Marco G5 Ultra	attached	n/a	On Stand
OPVI	6437	2 Stool, Backrest, Adj.Ht., On Casters (Is foot ring required)	floor	15"diameter x 19-27" ht	None
OPVI	6445	1 Tonometer, Applanation with 2 Prisms	stand attached	N/A	None
OPVI	6445a	1 Tonometer Mount	attached	N/A	None
OPVI	6448	1 Frame, Trial Lens, Adult, Light Weight, Metal, Holds 5 pairs	stored	n/a	None
OPVI	6449	1 Trial Lens Set, Minus cylinders, full diameter, 191 lenses:	stored	n/a	None
OPOI	6452	1 Monitor, Tel Screen Digital Imaging System, with out the slit lamp: software: Wall Mount: BACKING REQ	wall	TBD	E 115V: Data-Com
OPVI	6481	1 Lens, Volk, 20D Large, aspheric, in case	stored	n/a	None
OPVI	6482	1 Lens, Volk, 60D, Large, aspheric	stored	n/a	None
OPVI	6484	1 Lens, Volk, Superfield	stored	stored	Volk
OPVI	6487	1 Lens, 3 mirror Diagnostic: Gonio	stored	N/A	None
OPVI	6495	1 Kit, Foreign Body and Therapeutic Kit, Potaznick	stored	n/a	None
OPVI	6496	1 Forceps, Jeweler	stored	n/a	None
OPVI	9127	1 Sphygmomanometer, Aneroid, Hand Use, Portable, With Assortment	stored	n/a-stored item	None

ROOM NO 01-10-08
RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

ROOM NO 01-10-09
RMTITLE EXAM, EYE (EXISTING EQUIP TO TX-MEP SAME AS 10-03)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1753	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EYE CARE REFRACTION ROOM: SEE BACKING	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	6500	1	EXISTING COMPLETE EYE EXAM ROOM EQUIPMENT TO TRANSFER	VARIES	VARIES	MEP SEE RM 10-02

ROOM NO 01-10-10
RMTITLE OFFICE, PROVIDER

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	30"w x15"d x54"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-10-11

RMTITLE OFFICE, PROVIDER

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	30"w x15"d x54"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-10-12

RMTITLE OPEN WORK, TEAM SPACE (4 STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	4	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	4	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1406	1	Common Work Surface/Table Top Suitable for Function of Room/will have storage underneath	floor	72"lx36"d x42"h	None
OPVI	1500	4	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	4	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3061	1 Wastebasket, Office, 16 Gallon Slim Jim w/ Venting Channels Pack of 4	floor	22"w x11"l x25"h	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	1 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	3 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPOI	3916	1 Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com

DEPARTMENT TITLE: WIC**ROOM NO** 01-11-01**RMTITLE** TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-11-02**RMTITLE** BREASTFEEDING ROOM (REQ LIGHT DIMMER)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1051	1	Clock, Battery Operated-Atomic, Size Appropriate to space	wall	11.75" Diameter	Battery Op
OPVI	1160	1	Recliner, Patient, with Glider Option; Wall Saver, Clean Out,	floor	27.75"w x 29.5"d x 40.5"h	None
OPVI	1205	1	Lamp, Table, Decorator, 3 level bulb	desktop	15"w x15"d x 29"h	E 115V
OPVI	1410a	1	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	1755	1	Casework, Modular, Designed to Meet Space Req: BREAST FEEDING ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	2294	1	Diaper Pail, with Step	floor	12"w x 10"d x 32H	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
CPCI	3135	1	Diaper Changing Station, Horizontal Drop Down: TO BE PROVIDED & INSTALLED BY CONTRACTOR;	wall	35.25"L X3.75"D X22.25"H	None

ROOM NO 01-11-03

RMTITLE OFFICE, WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

ROOM NO 01-11-04

RMTITLE OFFICE, WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com

OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

ROOM NO 01-11-05
RMTITLE STORAGE, WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18'd x65'h	None
OPVI	2254	6	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-11-06
RMTITLE OFFICE, WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	

OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

ROOM NO 01-11-07

RMTITLE OFFICE, WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3914x	1	Printer, Multifunction: EXISTING EQUIPMENT TO TRANSFER	countertop	15"w x 18"d x 16"h	E 115V: Data/Com

ROOM NO 01-11-08

RMTITLE EDUCATION, HEALTH

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	2	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1005	3	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None

OPVI	1007	1 Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVICI	1500c	1 Workstation,Open Office Furniture System,Modular, 2 person	floor	Various	E 115V Quad: Data Com
OPOI	2941x	2 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	2 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	2 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

DEPARTMENT TITLE: SPECIALTY CARE**ROOM NO** 01-12-00**RMTITLE** MEDICATION ROOM (FOR PHN/WIC/AND OTHERS)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	5018	1	Sink, Smart for drug waste containment	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPSVI	5030b	1	Controlled Dispensing System - 2 Drawer Main-Countertop Unit (LEASE ITEM?)	Countertop	23"W x27"D x29"H	E 115V, Data-Com x 2
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	29"w x 79.6h"h x 34.2"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 32"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data

ROOM NO 01-12-01**RMTITLE** EXAM, SPECIALTY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None

OPVI	1008a	1 Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1 Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1 Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1 Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1 Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1 Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	0 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPVI	8572	1 Light, Examination,Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated

OPOI	8777	1	Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312	1	Table, Patient Examination, Powered (Midmark 626) 32" width seat with built in I.Q. Scale System	floor	32"w x 60"-76"l x 15.5"-37"h	E 115VAC
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO 01-12-02

RMTITLE EXAM, SPECIALTY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None

OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1	Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	0	Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1	Computer, Laptop, To be used with MOBILE Workstation: SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1	Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x 8"l x 10"h	None
OPVI	8572	1	Light, Examination, Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1	Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1	Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312a	1	Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer, recep	floor	28"w x 18"-37"h x 76" ext'd	E 115VAC, 50/60Hz, 8amps
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x 3.9w x 7"h	None
CPCI	9993	1	Track, Cubicle Privacy Curtain, of Appropriate Size & fabric type	ceiling	as required-see layout	None:

ROOM NO **01-12-03**
RMTITLE **EXAM, SPECIALTY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x 20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x 19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms, Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details

OPOI	2941	1 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1 Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1 Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1 Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1 Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1 Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1 Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1 Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3902	1 Workstation On Wheels; Securable for Laptop:	floor	18"w x 20"d x 42"h	E 115V for Charging
OPOI	3903	0 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3905	1 Computer, Laptop, To be used with MOBILE Workstation:SPECIFIED & PROVIDED BY OWNER	stored	TBD	E 115V; Wireless Data Com
OPVI	8301	1 Flag Set, Coded, 6 colors, (Located Exterior to Room)	wall	8"w x8"l x10"h	None
OPVI	8572	1 Light, Examination,Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPVI	8573	1 Light Source for Vaginal Speculum with Transformer (stored item)	countertop	6" l x 4"w x 4"h	E 120V for charger or Battery Operated
OPOI	8777	1 Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V
OPVI	9312a	1 Table, Patient Examination, Powered, Lowers to 14" w/clean assist roller sys, heated drawer, recep	floor	28"w x18"-37"h x76" ext'd	E 115VAC,50/60Hz, 8amps
OPVI	9854	1 Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None

CPCI 9993 1 Track, Cubicle Privacy Curtain, of ceiling as required-see layout None:
Appropriate Size & fabric type

ROOM NO 01-12-04

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	5	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-12-05

RMTITLE OFFICE, SPECIALTY PHYSICIAN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-12-06

RMTITLE OFFICE, SPECIALTY PHYSICIAN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None

OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO **01-12-07**
RMTITLE **OPEN WORK, TEAM SPACE (6 STAFF)**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	6	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	6	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1406	1	Common Work Surface/Table Top Suitable for Function of Room/will have storage underneath	floor	72"lx36"d x42"h	None
OPVI	1500	6	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	6	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	6	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903	6	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	6	Keyboard Tray	under desk	22" work track	None

DEPARTMENT TITLE: PUBLIC HEALTH**ROOM NO** 01-13-00**RMTITLE** RECEPTION, SHARED WITH WIC

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0152	1	Chair, Rotary, Bariatric, Std.Back, On Casters, Adjustable Arms	floor	28"w x21"d x40.5-45.5"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907x	1	Computer & Dual Monitor, Existing Equipment To Transfer, Mounting Option TBD	desk top	TBD	E 115; Data-Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPOI	3936a	1	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data
OPVI	3940	1	Keyboard Tray	under desk	22" work track	None

ROOM NO 01-13-02**RMTITLE** ALCOVE, WEIGHT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1758	1	Casework, Modular, Designed to Meet Space Req: WEIGHT ALCOVE: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details

OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVI	8915	1	Scale,Multipurpose, Stand On, Bariatric, Low Profile, With Stadiometer	floor	24"w x26"d 46"h	E 115V; or Batter Operated
OPVI	9310	1	Table, Pediatric Exam, Digital Weight with Measurement, Storage underneath	floor	42"l x20.8"d x41.9	E115V;

ROOM NO 01-13-03

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-13-04

RMTITLE OFFICE, INFECTION CONTROL

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

OPOI	3914	1	Printer, Multifunction, LaserJet mf copy/scan/fax	countertop	15"w x 18"d x 16"h	E 115V: Data/Com
------	------	---	---	------------	--------------------	------------------

ROOM NO 01-13-05

RMTITLE EXAM, PHN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1750	1	Casework, Modular, Designed to Meet Space Req: TYPICAL EXAM ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	8572	1	Light, Examination,Halogen, Table Mounted	wall or table	6x6x30	E 115V
OPOI	8777	1	Otoscope-Ophthalmoscope Wall Unit with Speculum Cover Dispenser	wall mount	7.1"w x 9.2"h x 3.9"d	E 115V

OPVI	9312x	1	Table, Patient Examination, Powered EXISTING EQUIPMENT TO TRANSFER	floor	TBD	E 115VAC
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None

ROOM NO 01-13-06

RMTITLE PHN MEDICATION ADMINISTRATION (MIDMARK CABINETRY?)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1008	1	Chair, Patient, 4 legs, with arms, Bariatric, with vinyl upholstery, Midmark, wt capacity 650 lbs	floor	30"w x20"d x seat ht 18"	None
OPVI	1008a	1	Chair, Patient, 4 legs, no arms, vinyl upholstery, Midmark, wt capacity 400 lbs	floor	18.5"w x19.5"d x seat ht 18"	None
OPVI	1334	1	Stool, Backrest, Footring, Adj.Ht., No arms,Casters	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1419	1	Table, Rectangular, On Casters, Optional Flip Top, Optional Drawer	floor	30"w x 24"d x 30"h	None
OPVI	1751	1	Casework, Modular, Designed to Meet Space Req: TYPICAL MED ROOM: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3041	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 8 gallon	floor	16.25"w x15.75"d x17.125"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3089	1	Dispenser, Quad Glove Box: CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	22.25"w x 10"h x 3.75"d	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None

OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5050	1	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	29"w x 79.6h"h x 34.2"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5054	1	Freezer, Undercounter:ADA:Scientific;- 15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 32"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPVI	9854	1	Sundry Jar Set of 5, for Exam Room Supplys (tongue depressors, alcohol wipes, cotton balls, etc...)	countertop	4.2l x3.9w x7"h	None

ROOM NO 01-13-07

RMTITLE OFFICE, PHN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5- Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37- 42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data- Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-13-08

RMTITLE OPEN WORK AREA, PHN (3 STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	3	Chair, Rotary,Std.Back,5- Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37- 42"h	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None

OPOI	2941	3 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	3 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3940	3 Keyboard Tray	under desk	22" work track	None

ROOM NO 01-13-09
RMTITLE ALCOVE, COPY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm

ROOM NO 01-13-10
RMTITLE STORAGE, PHN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0680	2	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18"d x65"h	None
OPVI	2254	4	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-13-11
RMTITLE STORAGE, PHN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	1	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None

OPVI	2403	2	Cart, Stainless Steel, 3 Shelf, Heavy Duty, 1000 lb capacity	floor	22 3/4"w x 49 3/8"l x 55"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 01-13-12

RMTITLE OPEN WORK AREA (4 STAFF)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	4	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	4	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1500	3	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	2	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPOI	2941x	2	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	4	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903	2	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3903x	2	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3916	1	Printer, Multi Function, Fax/Copy/Scan	countertop	24.5"w x 19"d x 25.5"h	E 115V:Data Com
OPVI	3940	4	Keyboard Tray	under desk	22" work track	None

DEPARTMENT TITLE: COMMUNITY SPACES**ROOM NO 01-05-02****RMTITLE STORAGE, KITCHEN (PREVIOUSLY TANK STORAGE)**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3040a	1	Wastebasket, Step-On, Plastic, Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None

ROOM NO 01-14-01**RMTITLE CONFERENCE ROOM; LARGER OF THE 2 MTG SPACES**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0212	40	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	22"w x20"d x32"h	None
OPVI	1402a	20	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx 24"d x30"h	E 115V-floor: Data-Com
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4112	1	Projector, Ceiling Mount, wireless capability TO BE SPECIFIED BY OWNER: STRUCTURAL	ceiling	13"w x 10.5"d x 4.8"h	E 115V; Data-Com
OPCI	4120b	1	Screen, Projection, Powered, HDTV Format; Concealed in Ceiling: STRUCTURAL CONSIDERATIONS	wall/ceiling	139"w x8"d x78"h	E 120V, Control Switch
OPVI	4123	1	Board, Smart; Portable on Casters	FLOOR	75" panel	None
OPOI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x5.12"h x2.6"d	POE
OPOI	8202	1	Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1	Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated

ROOM NO 01-14-02**RMTITLE CONFERENCE ROOM (SMALLER OF THE 2 MTG SPACES)**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0212	30	Chair, Conference Type, Stacking, Upholstered, for high density area; high wt capacity	floor	22"w x20"d x32"h	None

OPVI	1402a	15	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"lx 24"d x30"h	E 115V-floor: Data-Com
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4112	1	Projector, Ceiling Mount, wireless capability TO BE SPECIFIED BY OWNER: STRUCTURAL	ceiling	13"w x 10.5"d x 4.8"h	E 115V; Data-Com
OPCI	4120b	1	Screen, Projection, Powered, HDTV Format; Concealed in Ceiling: STRUCTURAL CONSIDERATIONS	wall/ceiling	139"w x8"d x78"h	E 120V, Control Switch
OPOI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x5.12"h x2.6"d	POE
OPVI	4151	1	Lectern, For Education Environments	floor	30"w x 42"h x 24"d	None

ROOM NO 01-14-04

RMTITLE KITCHEN, LIGHT COMMERCIAL, WARMING KITCHEN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3047a	1	Trash Collector w/Lid,and Dolly, 44 gallons (color TBD by Function)	floor	24"diameter x 37.5"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
CPCI	3159	1	Sink, Handwashing, w/Eye Wash station	wall	10"d x 15"w x 5"h	H/C water; drain
CPCI	3160	1	Sink, 3 Well, SS, Drop In, Deep Bowl, Goose Neck Faucet	drop in sink	56"w x 25"d x 12"h	H & C Water, Drain
CPCI	3192	1	Hood Exhaust with Required Fire Suppression for cooking areas; optional CFM	wall mount	TBD	TBD
OPVI	3215	1	Coffeemaker, Automatic, Air Pot Twin Brewing System	countertop	17.75"w x17.25"d x22"h	E 120/240V Dedicated; 4 Wire;17 amp; Water Supply
OPVI	3216b	1	Rack, Coffee Container for Air Pots, with drip trays	stored	20.5"w x 11.5"d x 15"h	None

OPVI	3216c	1 Pots, Air; Glass lined; 2.2 liter; lever action; hot up to 8 hrs (pkg 6 qty per case)	stored	brew ht 12.25"h/lid closed 14 9/16" h	None
OPSVCI	3250	1 Ice Station, Floor Model, Bulk: With Optional Hotel Ice Dispensary	floor	22"w x31"d x 75.5"h	See Mfg Details
OPVI	3255	2 Microwave Oven,Commercial	overshelf	22"w x 19"d x 14"h	E 120V: 15 amp: NEMA 5-15
OPCI	3266	1 Range, Free Standing, Electric, Dual Oven Compartments	floor	29 15/16"w 26 1/8"d x 47 1/8"h	E Ded 120-240V;60Hz; 40-50 Amp
OPVI	3268	1 Refrigerator, Freezer: Commercial, SS, 2 Doors, on casters, 6 shelves, 1HP, exterior temp display	floor	54 1/8"w 29.5"d x84"h	E Ded 115V; 9.6 amps;NEMA 5-15P
OPCI	3352b	1 Dishwasher, Automatic, Undercounter, Commercial Unit; Reaches 185 degree water temp on final rinse	floor	33"h x 24"w x 24"d; See Details	E 3AC-208V;3x20 Amp;NEMA L15-20;1 Hot,2 Cold;drain

DEPARTMENT TITLE: PHARMACY**ROOM NO 01-15-00****RMTITLE WAITING, PHARMACY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1007	7	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVI	1019	22	Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1024	2	Seating, 3 Person, Couch Type	floor	74"w x29"d x30.75"h	None
OPVI	1026	4	Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1302	2	Rack, Magazine/Pamphlets,Wall Mounted 1 Pocket	wall	14" w x4"d x14.5"h	None: Consider Backing
OPVI	1410a	11	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	3043	2	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4132	2	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	37.8"w x 22.1"h x2.2"d	E 115V; cabling; Data
OPVI	5010	1	Medication Disposal System; Drop Box; ??? 18 or 38 Gal; for Schedules II-V; Secures to Wall or Floor	floor	22.5"w x 22.5"l x 43.75"h	None
OPSVI	5048	1	Kiosk, Script Pro: FUTURE CONSIDERATION (for waiting area)	TBD	TBD	E 115V Ded; Data-Com
OPSVI	5049	2	Monitor, Status Message System: Bingo Boards, Script Pro: BACKING REQUIRED	wall mount	36.2"w x21"h x2.12"d	E 115V Ded; Data-Com

ROOM NO 01-15-01**RMTITLE RECEPTION, PHARMACY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0150x	2	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	approx 24.5" x28" x37"	None

OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPOI	2941x	2	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPOI	3903	3	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPVI	3940	3	Keyboard Tray	under desk	22" work track	None
OPSVI	5044	1	Signature Pad, Electronic: Integrated PC via USB port; Specified by Owner	countertop mt	7.5"w x 4.75"h x7.5"d	None
OPSVI	5044x	2	Signature Pad, Electronic: Integrated PC via USB port; EXISTING EQUIPMENT TO TX	countertop mt	7.5"w x 4.75"h x7.5"d	None

ROOM NO **01-15-02**
RMTITLE **OFFICE, CONSULT**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None

OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO **01-15-03**
RMTITLE **OFFICE, CONSULT**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO **01-15-04**
RMTITLE **OFFICE, HC CONSULTANT**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None

OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO 01-15-05
RMTITLE OFFICE, CONSULT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPSVI	5042x	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt: EXISTING EQUIP TO TRANSFER	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO 01-15-06
RMTITLE OFFICE, CONSULT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None

OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO 01-15-07

RMTITLE OFFICE, CONSULT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5042	1	Checkpoint, SP: Interactive RX verification, tracking, mgmt	countertop	15.5"w x 22"h x 10"d	E115V;50/60hz;1.5amp;d ata/com

ROOM NO 01-15-08

RMTITLE BREAK ROOM, STAFF *

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1005	1	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp;Water Supply;NEMA 5-15P
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	32.6" x 35"d x 70"h	E 115V; Water Supply for Icemaker

OPVI 3349 1 Microwave Oven,Household Type, 2.2 countertop 13.5"h x24"w x 18.5"d E 120V: 13.9 amp
Cu Ft, 1100 watt

ROOM NO 01-15-09

RMTITLE TOILET, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-15-10

RMTITLE STORAGE, RX *

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
------	------	---	---	-------	------------------	------

ROOM NO 01-15-11

RMTITLE BREAK OUT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

CPCI	3900	1	Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com
------	------	---	--	------	-----	----------

ROOM NO 01-15-12

RMTITLE OFFICE, SUPERVISOR *

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150x	1	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	approx 24.5" x28" x37"	None

OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
------	------	---	---	-------	------------------	------

OPVI	0700x	1	Bookcase/Lateral file combination unit: EXISTING EQUIPMENT TO TX	floor	36x18x64	none
------	-------	---	--	-------	----------	------

OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
------	-------	---	---	---------	------------------	----------

OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
------	-------	---	--	-------	------------------	------

CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-15-13

RMTITLE STORAGE, RX

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	1470	1	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design

ROOM NO 01-15-14

RMTITLE PHARMACY AREA, MAIN

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150x	6	Chair, Task; Rotary, Std.Back: EXISTING EQUIPMENT TO TRANSFER	floor	approx 24.5" x28" x37"	None
OPVI	0181	1	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPVI	0181x	6	Chair, Rotary,Std.Back:Extended Height: EXISTING EQUIPMENT TO TRANSFER	floor	approx 24.5"w x28"d x48"h	None
OPSVI	1470	3	Furnishing, Cabinetry, Shelving For Pharmaceutical Areas-Per RC Smith	floor/walls	Per Design	Per Design
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPOI	2941x	5	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	5	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	3	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062x	4	Wastebasket, Office,Tall, 23 gal, with flip lid (Blue Recycling-HIPPA): EXISTING EQUIPMENT TO TX	floor	20lx11wx30h	None
OPCI	3125	1	Rack, Coat, Size for up to 12 hooks for staff: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	REFER TO DESIGN PKG	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm

CPCI	3900	6 Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com
OPOI	3903	5 Computer, Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3903x	2 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913x	1 Printer, Laser, Individual EXISTING EQUIPMENT TO TX	countertop	TBD	E 115V; Data Comm
OPCI	5006	1 Filtered Water System, Digital, Fillmaster with 4 satellite stations	wall-undercount	varying sizes see mfg details	E 115V; Water Supply
OPOI	5018x	4 Sink, Smart for drug waste containment: EXISTING EQUIPMENT TO TX	countertop	15.5"w x 12.5"h x 9.5"d	E 115V or Battery
OPOI	5021	2 Counter, Tablet-Kirby Lester; with built in scanning: EXISTING EQUIPMENT TO TRANSFER	countertop	8"w x 12.5"d x 14.6"	E 110V
OPSVI	5027	1 Pharmacy, Robotic Medication Dispensing, EXISTING TO BE TRANSFERRED & COORDINATED BY	floor	152x35x76.5	See Scriptpro Design
OPOI	5043	1 Printer, Thermal, Label; (sits on ScriptPro System) EXISTING EQUIPMENT TO TX:	countertop	5.5"d x 14.5"h x 12"w	E115V;50/60hz;2amp;dat a-com
OPOI	5043x	1 Printer, Thermal, Label; (sits on ScriptPro System) EXISTING EQUIPMENT TO TX:	countertop	5.5"d x 14.5"h x 12"w	E115V;50/60hz;2amp;dat a-com
OPSVI	5050	1 Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	29"w x 79.6h"h x 34.2"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	5051x	1 Refrigerator, Scientific;Double door; EXISTING EQUIPMENT TO TX	floor	60"w x 80"h x 30"d	E-Ded 115V; 11.5 FLA;data
OPSVI	5054	1 Freezer, Undercounter:ADA:Scientific;-15 to-30 Deg C;Onboard monitoring sys	floor	24"w x 32"h x 26.75"d	E-Ded 115V; 5.75 FLA; Data
OPSVI	5055	1 Refrigerator, Undercounter:ADA:Scientific;2-10 deg C;Onboard monitoring sys	floor	24.3"w x 31.9"h x 26.75"d	E-Ded 115V; 5 FLA;data

ROOM NO 01-15-15

RMTITLE WAITING, PHARMACY

GROUP **EQNO** **QTY** **DESCRIPTION** **LOCATION** **SIZE** **UTILITIES**

OPVI	1007	2	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x 22"d x 33"h	None
------	------	---	--	-------	---------------------------	------

OPVI	1019	2 Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1026	4 Seating, 1 Person, Upholstered	floor	30"w x29"d x30.75"h	None
OPVI	1410a	2 Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	1421	1 Table, Sofa Suitable for Lounge/Waiting Areas	floor	30" diameter x 22"h	None
OPSVI	5049x	1 Monitor, Status Message System: Bingo Boards, Script Pro: BACKING REQ;EXISTING EQUIPMENT TO TX	TBD	TBD	E 115V Ded; Data-Com

DEPARTMENT TITLE: ADMINISTRATION**ROOM NO 01-16-00****RMTITLE WAITING AREA, ADMINISTRATION**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1007	1	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVI	1019	2	Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1410a	1	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4132	1	TV, Monitor Smart LED, 42" WITH WALL MOUNT: OWNER TO SPECIFY: REQUIRES BACKING	wall mount	37.8"w x 22.1"h x2.2"d	E 115V; cabling; Data

ROOM NO 01-16-01**RMTITLE RECEPTION, ADMINISTRATION**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0151	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms, High Back	floor	27"w x 18"d x 40-45"h	None
OPVI	0430d	1	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1480	1	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-16-02**RMTITLE COPY/WORK ROOM**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
--------------	-------------	------------	--------------------	-----------------	-------------	------------------

OPVI	1405	2	Mail Cubbies (Risers) Optional 24 or 36 mail slots (REQUIRES BACKING)	wall mounted	34"w x 12"d x15"h	None
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x48"h	None
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm
OPVI	3780	1	Meter, Postage, Automatic: High Productivity Unit;EXISTING EQUIPMENT TO PLACE ON	countertop	65"l x 25"d x 26"h	E 120V; Data/Com x 2
CPCI	3900	1	Data-Com, Spare for future use (to be located per instruction)	wall	N/A	Data-Com

ROOM NO 01-16-03

RMTITLE TOILET, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 01-16-04

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0680	2	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18"d x65"h	None
OPVI	2254	3	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None

ROOM NO 01-16-05

RMTITLE OFFICE, ADMINISTRATOR (PANIC BUTTON)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	4	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none

OPVI	1412	1	Table, Round, Diameter, 35"	floor	35" diameter x29"h	None
OPVI	1501	1	Workstation, Office Furniture System, Sit To Stand	floor	See Design Details	E 115V Quad; Data; Com
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPCI	4131	1	Television, Flat Panel, 65"; LED; Smart TV; with mount; owner to coordinate: REQUIRES BACKING	wall/ceiling mt	57.6"w x 2.5"d x 33.2"h	E 115V; Cabling; Data

ROOM NO 01-16-06

RMTITLE KITCHENETTE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1005	4	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVI	1402a	1	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"l x 24"d x30"h	E 115V-floor: Data-Com
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp;Water Supply;NEMA 5-15P
OPVI	3269d	1	Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	32.6" x 35"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1	Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain

OPVI 3349 1 Microwave Oven,Household Type, 2.2 countertop 13.5"h x24"w x 18.5"d E 120V: 13.9 amp
Cu Ft, 1100 watt

ROOM NO 01-16-07

RMTITLE CONFERENCE ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0211	12	Chair, Conference Type, Highback, On Casters, w/arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	1402	1	Table, Formal Conference Shape & Preferred Base Type TBD; (phone & computers)	floor	60"w x168"l	E 115V-floor; Data-Com
OPOI	2945	1	Telephone, Conference: TO BE SPECIFIED BY OWNER	desktop	14X13X3	E 115V; Data-Com
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3310	1	Refrigerator, Undercounter, ADA 5.5 cubit ft, auto defrost, SS	floor	23.63"w x 23.63"d x 32"h	E 115V
OPCI	4131	1	Television, Flat Panel, 65"; LED; Smart TV; with mount; owner to coordinate: REQUIRES BACKING	wall/ceiling mt	57.6"w x 2.5"d x 33.2"h	E 115V; Cabling; Data
OPOI	4149	1	Scheduling System for Conference Room areas; 7" monitor display to be mtd outside of mtg space	wall mount	8.07"w x5.12"h x2.6"d	POE

ROOM NO 01-16-08

RMTITLE OFFICE, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm

OPOI	2941x	1 Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1 Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-16-09

RMTITLE OFFICE, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 01-16-10

RMTITLE OFFICE, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None

OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com

ROOM NO 01-16-12

RMTITLE OFFICE, SECURITY (BY PRIMARY CARE)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	1005	2	Chair, Side, No Arms, Vinyl Upholstery, Stackable	floor	18"w x19"d x32"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941x	1	Telephone, Standard: EXISTING EQUIPMENT TO TRANSFER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903x	1	Computer, Personal, Complete with Monitor, Keyboard, etc: EXISTING EQUIPMENT TO TX	desktop	TBD	E 115V; Data Com
OPOI	3913	1	Printer, Laser, Individual B & W, Size & PPM TBD BY OWNER	countertop	16.3"w x16.6"d x17.7"h	E 115V; Data Comm
OPCI	4136	2	Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

DEPARTMENT TITLE: DENTAL**ROOM NO 02-17-00****RMTITLE WAITING, DENTAL**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1007	8	Chair, Patient, 4 legs, with arms, Bariatric	floor	approx 31"w x22"d x 33"h	None
OPVI	1019	28	Chair, Tandom, 1/2/3 seats w/various arm arrangements	floor	23.5"w x 24"d x32"h	None
OPVI	1024	2	Seating, 3 Person, Couch Type	floor	74"w x29"d x30.75"h	None
OPVI	1302a	2	Rack, Magazine/Pamphlets,Wall Mounted 3 Pocket	wall	14" w x4"d x23"h	None: Consider Backing
OPVI	1410a	12	Table, End, Drum Round 24" Diameter	floor	24" Diameter x22"h	None
OPVI	1421	1	Table, Sofa Suitable for Lounge/Waiting Areas	floor	30" diameter x 22"h	None
OPVI	3043	3	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPCI	4136	2	Television, Flat Panel, 55" Wall Mount: OWNER TO SPECIFY: REQUIRES BACKING FOR SUPPORT	wall mount	49.2"w x 29"h x 3.1"d	E 115V; Cabling; Data

ROOM NO 02-17-01**RMTITLE RECEPTION, DENTAL**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	3	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0430d	3	Cabinet, Filing,Lateral, Mobile Pedestal, Box/File	floor	20"d x 15"w x22"h	None
OPVI	1480	3	Pneumatic Sit-to-Stand Adjustable Single Column Riser Desk Converter	desktop	30.7"W X 6.9-17.9" ht adjustability	None
OPOI	2941	3	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	3	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None

OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPVI	3760	1	Copier,Medium Volume, 40 PPM; Multi Function,w/Stand: EXISTING EQUIPMENT TO TX	floor	51"w x 26"d x 45"h	E-Ded 115V NEMA 5-15R
OPOI	3903	3	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3936a	3	Scanner, Desktop, Auto document feeder, Duplex, Color: EXISTING EQUIPMENT TO TX	desktop	11.5"w x 6.2"d x 6.6"h	E 110V; Data

ROOM NO 02-17-02

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-03

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-04

RMTITLE CONFERENCE ROOM-BREAK ROOM

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0180	24	Chair, Rotary,Std.Back,4-Legs, Nesting,Ergonomic,NO Arms	floor	22"w x20"d x32"h	None
OPVI	1402a	4	Table, Educational Type, Flip Top, Nesting, w/Optional Casters	floor	60"l x 24"d x30"h	E 115V-floor: Data-Com
OPVI	2524	1	Board, Marker, Cleanable, OPTIONAL AS PORTABLE	wall	36"w x 72"h	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3043	1	Wastebasket, Fiberglass, Half Moon Shape 12 gal, Open Top; Textured Black Powder Coat; Perforated	floor	18"w x9"d x32"h	None
OPVICI	3216	1	Coffeemaker,Automatic, Low Profile, 3 warming Burners	countertop	16.5"w x 17.75"d x 17"h	E 120V Dedicated; 15 amp;Water Supply;NEMA 5-15P

OPVI	3269d	2 Refrigerator, 2 door, freezer in lower section, 24 Cu Ft, w/icemaker, SS	floor	32.6" x 35"d x 70"h	E 115V; Water Supply for Icemaker
OPSVCI	3307	1 Ice Maker, Countertop, Nugge Ice/Water Disp, air cooled, 20# ice bin, up to 250# ice in 24 hr period	countertop	16.25"w x24"d x42"h (includes 6" legs)	E 115V, Ded 15 Amp Brkr, H2O, see details for Drain
OPVI	3349	1 Microwave Oven, Household Type, 2.2 Cu Ft, 1100 watt	countertop	13.5"h x24"w x 18.5"d	E 120V: 13.9 amp
OPCI	4131	1 Television, Flat Panel, 65"; LED; Smart TV; with mount; owner to coordinate: REQUIRES BACKING	wall/ceiling mt	57.6"w x 2.5"d x 33.2"h	E 115V; Cabling; Data

ROOM NO 02-17-05

RMTITLE TOILET, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-06

RMTITLE TOILET, STAFF

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-08

RMTITLE PUMP ROOM (AIR & VAC SYSTEM TO USE EXISTING EQUIP)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVCI	9901	1	Compressor, Dental Air: with switch & intake kit: EXISTING TO TRANSFER: DETAILS BY BENCO	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVCI	9902	2	Vacuum Pump System-Dry Vac: w/remote control panel 2 switch: EXISTING TO TRANSFER: DETAILS	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVI	9903	2	Separator, Almagam for 10 to 20 Users: By Benco: Midmark HG5 HV: EXISTING TO TX: DETAILS BY BENCO	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS
OPSVI	9904	2	Reverse Osmosis Universal Dental Water Supply; High Efficiency: EXISTING TO TX: DETAILS BY BENCO	floor	PER DESIGN DRWGS	SEE DESIGN DETAILS

ROOM NO 02-17-09

RMTITLE STORAGE, CYLINDER TANK

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	2425	1	Rack, Cylinder tank, to hold 6 cylinders; SPECS & DETAILS BY BENCO; WILL REQUIRE WALL BACKING	wall mount	30"w x 30"d x 36"h	NONE

ROOM NO **02-17-10**
RMTITLE **OPERATORY, CLOSED (Microscopic Procedures RM)**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET MT, REQ BACKING	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg Details	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVCI	6004	1	Microscope, Dental Procedure; A Series; SEE MFG DETAILS; ceiling mount structure to support 500 lbs	ceiling	4.70"w x3.25"h x3.0"d + column length	PER BENCO DRWGS
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

OPVI	8700	1 Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-11

RMTITLE OPERATORY, CLOSED

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET MT, REQ BACKING	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg Details	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS

OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-12

RMTITLE OPEN OPERATORY BAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6007	1	3 O'Clock Upper Storage, 3 door unit Wall mounted; Open Bay With Owner Preferred Dispensers-ADEC	wall	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-13
RMTITLE OPEN OPERATORY BAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-14
RMTITLE OPEN OPERATORY BAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

OPCI	4133	1 TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1 Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1 12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1 Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1 Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1 Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1 Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1 Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1 Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1 Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-15

RMTITLE OPEN OPERATORY BAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None

OPVI	5706a	1 Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1 12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1 Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1 Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1 Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1 Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO **02-17-16**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details

OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO **02-17-17**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS

OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14
ROOM NO	02-17-18					
RMTITLE	OPEN OPERATORY BAY					
<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO **02-17-19**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320- 1 C14

ROOM NO **02-17-20**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS

OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6011	1	Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6047	1	Light, Dental Operating, Single Head, Console Mounted	floor/stored		See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO **02-17-21**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS

OPSVCI	6011	1 Console, Central W/End Sink : storage for Radiographic Unit & Supplies, ETC...: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1 Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1 Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1 Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO **02-17-22**
RMTITLE **OPEN OPERATORY BAY**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133	1	TV, Flat Panel, LED, Healthcare Grade, 22" Diagonal screen: DENTAL OPEN OP BAYS	wall mount	24.6"w x 15.5"h x 1.6"d	E 115V; cabling; Data
OPSVI	5760	1	Radiography, Dental, Digital Intra Oral: PassThru CENTRAL CONSOLE; SEE MFG DETAILS	stored/wall	SEE MFG DETAILS	PER PLAN MECA/BENCO DRWGS
OPSVI	6008b	1	9 O'Clock, Accessory Console, 3 door unit Wall mounted;	floor	86x33x16	E, Data, H/C water
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6048	1	Light, Dental Operating, Single Head, Wall Mount	ceiling	6x48x24	See ADEC Details
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS

OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1 Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1 Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-24

RMTITLE LAB, DENTAL (CONFIRM PREFERRED CASEWORK SOURCE)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1338	3	Stool, Backrest, Adj.Ht with footring, casters optional, for standing height workstations	floor	Overall Width 25": Seat 16" Diameter	None
OPVI	1759	1	Casework, Modular, Designed to Meet Space Req: DENTAL DEPT: SEE BACKING DETAILS	floor/wall	See Mfg Design Details	See Mfg Details
OPVI	2519	1	Board,Chalkless,Cleanable Surface,w/Tray,	wall	48"w x1"d x48"h	None
OPOI	2944	1	Telephone, Wall Mounted: TO BE SPECIFIED BY OWNER	wall	8x3X10	Comm.
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPSVI	3100	1	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	1	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	6050	2	Macro Cab with Dust Collector: BY BENCO	countertop	12x12x15	E 115V
OPVI	6052	1	Mixer, Combination,Power/Vacuum,w/bench stand, complete: SEE BENCO DETAILS	countertop	8x8x12	E 115V
OPVI	6054	1	Bin,Plaster,3 Compartment, 75# capacity (REQUIRES SUPPORT BACKING) SEE BENCO DETAILS	wall	21x8x18	None
OPSVI	6058	2	Pot, Curing, Pressure: 8 quart: By BENCO	countertop	12" diameter x 10"h	E 115V

OPSVI 6059 2 Knife, Proform Electric for dental labs: countertop 8"w x 6"d x 3"h E 110V
By BENCO

ROOM NO 02-17-25

RMTITLE OFFICE, ASSISTANT SUPERVISOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com

ROOM NO 02-17-26

RMTITLE OPEN WORK AREA, TEAM SPACE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	8	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0385	1	Bookcase,Non-Sectional,Open, 5 Adj. Shelves	floor	30"w x15"d x54"h	None
OPVI	0433d	1	Cabinet, desk height, with doors, for storage of supplies, and worksurface for printers, etc...	floor	36"w x 20"d x26"h	None
OPVI	1500b	8	Workstation,Open Office Furniture System,Modular, Small Scale	floor	Various	E 115V Quad: Data-Com

OPOI	2941	8 Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	8 Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPVI	3039b	1 Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
CPCI	3115	1 Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3903	8 Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPOI	3929	1 Copier, Multifunction, 44 ppm, Color/B/W (stand optional)	countertop	22.3"w x 23.4"d x 24.2"h	E 115V; Data/Com

ROOM NO 02-17-28

RMTITLE OFFICE, DENTAL DIRECTOR

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	4	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None
OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1412	1	Table, Round, Diameter, 35"	floor	35" diameter x29"h	None
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907	1	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com

ROOM NO 02-17-29

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0680	1	Cabinet, Storage, 6 Adj.Shelves,2 Doors,Lock	floor	36"w x18"d x65'h	None
OPVI	2254	2	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 02-17-30

RMTITLE EVS CLOSET

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2609	1	Rack, Wall Mounted to hold mops, etc...	wall mount	34"w x 1.5"d x4.2 5"h	None
OPVI	3046	1	Trash Collector, 32 Gal Brute, Round, with lid, dolly & wrap around caddy for janitorial services	floor	22"diameter x37.5"h with dolly & lid	None

ROOM NO 02-17-32

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-33

RMTITLE TOILET, PATIENT

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3043b	1	Wastebasket, Round, Open Top, 12 gallon, with rigid liner, floor protective base, fire safe	floor	15" diameter x 30"h	None

ROOM NO 02-17-34

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	6	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	2	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None
OPVI	2254b	2	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x72"h	None

OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
------	------	---	---	-------	------------------	------

ROOM NO **02-17-35**
RMTITLE **OPERATORY, CLOSED**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET MT, REQ BACKING	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg Details	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-37

RMTITLE PANORAMIC XRAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1417	1	Table/Stand, For Computer, Printer, or opt paper storage below, On Casters	floor	30"w x 24"d x 29"h	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5761a	1	Radiographic, Dental, Panoramic-Digital:REQUIRES BACKING: EXISTING EQUIP TO TX	floor	See MFG Details	SEE PLAN MECA DETAILS
OPVI	5768a	1	Rack, Apron; Single; BACKING REQUIRED: BY BENCO	wall-Backing	3"H • 5"L prongs; on 16" centers	PER BENCO DRWG
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

ROOM NO 02-17-38

RMTITLE STORAGE

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	1	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None
OPVI	2254a	1	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x69"h	None

ROOM NO 02-17-39

RMTITLE OFFICE, MANAGER

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0150	1	Chair, Rotary,Std.Back,5-Legged,Ergonomic,Adj.Short Arms	floor	27"w x18.5"d x 37-42"h	None
OPVI	0200	2	Chair, Fixed Back & Seat, Arms, No Casters, Desk Side Chair	floor	22"w x20"d x32"h	None

OPVI	0700	1	Bookcase/Lateral file combination unit	floor	30"w x 20"d x 66"h	none
OPVI	1500	1	Workstation, Furniture System, Style, Size TBD Designer	floor	Various	E 115V Quads: Data-Comm
OPOI	2941	1	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
CPCI	3115	1	Robe Hook, Double: TO BE PROVIDED & INSTALLED BY CONTRACTOR	wall	6x6x3	
OPOI	3907	1	Computer & Dual Monitor, Mounting Option TBD	desk top	TBD	E 115; Data-Com

ROOM NO 02-17-40

RMTITLE LOCKERS AREA

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
CPCI	1095	1	Lockers, Type & Qty TBD: Either Built In or Mfg	floor	Per Design	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	9997	1	Bench, Sitting Suitable for Location, Padded Seat, ADA/Bariatric	floor	48" x20.25"d x18"h	None

ROOM NO 02-17-41

RMTITLE PANORAMIC XRAY

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	1417	1	Table/Stand, For Computer, Printer, or opt paper storage below, On Casters	floor	30"w x 24"d x 29"h	None
OPVI	3039a	1	Wastebasket, Rectangular, plastic, water/fire resistant,28qt	floor	14"l x11"w x16"h	None
OPOI	3087	1	Dispenser, Glove for Single Box, CLEAR (HORIZONTAL OR VERTICAL MOUNT)	wall mount	5.5"w x 3.75"d x 10"h	None
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	5761	1	Radiographic, Dental, Panoramic-Digital: WALL MOUNT REQUIRES BACKING SEE PLAN MECA DETAILS	floor	See MFG Details	SEE PLAN MECA DETAILS

OPVI	5768a	1 Rack, Apron; Single; BACKING REQUIRED: BY BENCO	wall-Backing	3"H • 5"L prongs; on 16" centers	PER BENCO DRWG
OPSVCI	6065	1 Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS

ROOM NO **02-17-42**

RMTITLE **CLEAN ROOM-ICC**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPSVI	3100	2	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	2	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPOI	5274	1	Incubator, Dry Heat, Open, 10/20 well, digital	countertop	11.3"w x 7.9"d x 3.9"h	E 115V
OPVI	6002	18	Amalgamator High Speed: BY BENCO	countertop	TBD	E 115V
OPSVI	6010a	1	Dental Cabinetry System: Clean Sterile: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPOI	6037	60	Handpiece, High Speed, Fiberoptic 5 hole: Impact Air 45:	stored	stored	None
OPSVI	6043	12	Intra Oral Camera-CS 1500: BY BENCO	attached	See Benco Details	None
OPSVI	6067	6	Sensor Kit Combo: MFG TBD; BENCO ITEM	stored	varies	None
OPSVI	6069	16	Sensor Kit Size 2: MFG TBD; BENCO ITEM	stored	varies	None
OPVI	6124	2	Cleaner/Lubricator, Handpiece, Air Driven, w/ handpiece: SEE BENCO DETAILS	countertop	7.5"w x 16.3"d x8.5"h	Air Supply 65 psi not to exceed 140 psi
OPVI	9221	2	Sterilizer, Table Top,Electric,Steam Type,w Printer & Pouch Rack: SEE MFG DETAILS	countertop	23.8"w x17.8"d x17.8"h	E Ded 115 VAC, 50/60 Hz, 15 A Single Phase

ROOM NO **02-17-43**
RMTITLE **SOILED ROOM-ICC**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	3040	1	Wastebasket, Step-On,Plastic,Rectangular, Red/Bio Hazard 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPVI	3040a	1	Wastebasket, Step-On,Plastic,Rectangular, 23 gallon	floor	19.75"w x 16.5"d x 32.63"h	None
OPSVI	3088	1	Container, Sharps: Wall Mountable or Countertop: Lockable: 5 QUART	wall mount	1.7513"w x 18"h x5.5"d	None
OPSVI	3100	2	Dispenser, Paper Towel: TYPE TBD BY OWNER BASED UPON CURRENT JANITORIAL SERVICE	wall mount	TBD	None
OPSVI	3102	2	Dispenser, Soap:Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall mount	7"w x4"d x11.5"h	None
OPSVI	3102a	1	Dispenser, Waterless Hand Sanitizer: Type TBD BY OWNER-CURRENT JANITORIAL SERVICE	wall	TBD	None
OPSVI	6010e	1	Dental Cabinetry System: Soiled Room Processing: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPVI	6021a	2	Disinfector, Thermal, for Dental Instruments, Undercounter: SEE BENCO DETAILS	undercounter	24x24x33.5	SEE BENCO DESIGN DRWGS
OPSVCI	6056	1	Autoclave:Single or Double Door-Pass Thru: CONTRACTOR TO REFER TO MFG DESIGN DETAILS	floor	30."w x40.25"d x 74.5"h	SEE MFG DETAILS
OPVI	6407	1	Cleaner, Ultrasonic; With Digital Timer (BENCO) Drain, Mounting Rim: By BENCO	Countertop	See Benco Details	See Benco Details

ROOM NO **02-17-44**
RMTITLE **OPEN WORK AREA, TECH**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPSVI	5050	2	Refrigerator, Scientific;Single door; 2-10 degree C;Onboard monitoring sys;	floor	29"w x 79.6h"h x 34.2"d	E-Ded 115V; 7.5 FLA;data-com
OPSVI	6010f	1	Dental Cabinetry System: Tech Workspace: Design by Midmark & Benco	floor	See Midmark-Benco Details	See Midmark-Benco Details

ROOM NO **02-17-46**
RMTITLE **STORAGE**

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	2254	3	Shelving Unit, Open Wire Shelves, Mobile, With Liner for Bottom Shelf	floor	36"w X21"d x69"h	None

OPVI	2254b	1	Shelving Unit, Open Wire Shelves, Mobile, w/drawer glide system & bins(optional solid bottom shelf)	floor	60"w x21"d x72"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None

ROOM NO 02-17-47

RMTITLE OPEN WORK AREA, TECH

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPVI	0181	4	Chair, Rotary,Std.Back,4-Legs, Ergonomic,NO Arms, Extended Height	floor	24.5"w x28"d x48"h	None
OPOI	2941	4	Telephone, Standard: TO BE SPECIFIED BY OWNER	desktop	10"D X 8"W X 4"H	Data-Com
OPVI	3039b	1	Wastebasket, HIPPA Compliant, with paper slot, locking top, 24 gallon	floor	15.25"w x 15.25"l x24.25"h	None
OPVI	3062	1	Wastebasket, Office,Tall, 23 gal, with flip lid (COLOR TBD)	floor	22"l x11"w x30"h	None
OPVI	3758	1	Copier, High Vol, 55ppm, MF, networkable: EQUIP TO LEASED?- SPECS TBD BY OWNER	floor	60"w x 30"d x 42"h	E V-TBD; Data-Comm
OPOI	3903	4	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPSVI	6010f	1	Dental Cabinetry System: Tech Workspace: Design by Midmark & Benco	floor	See Midmark-Benco Details	See Midmark-Benco Details

ROOM NO 02-17-48

RMTITLE OPERATORY, CLOSED

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET MT, REQ BACKING	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None
OPVI	5706a	1	Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None

OPSVI	5760a	1	Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg Details	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1	3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1	9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1	12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1	Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1	Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1	Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1	Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1	Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1	Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1	Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

ROOM NO 02-17-49

RMTITLE OPERATORY, CLOSED

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	3903	1	Computer,Personal & Monitor for Desk Use: TO BE SPECIFIED BY OWNER	desktop	24"w x 18"d x 22"h	E 115V; Data Com
OPCI	4133a	1	Mount, Wall, Swing out arm, Highly Articulating; REQUIRES BACKING	wall mount	12"w x 13"h x 1.5"d	None: Note utility location
OPCI	4134	1	Television, Flat Panel, LED, Healthcare Grade, 27";DENTAL CLOSED OPS; IF NOT CABINET MT, REQ BACKING	wall	38.1"w x 22.8"h x 4"d	E-Comm-Cable
OPVI	5706	1	Apron,XRay Protective,AdultDental Patient w/Collar: BY BENCO	stored	n/a	None

OPVI	5706a	1 Apron,XRay Protective,Child Dental Patient w/Collar: BY BENCO	stored	21x22	None
OPSVI	5760a	1 Radiography, Dental, Digital Intra Oral: WALL MOUNTED UNIT-REQUIRES BACKING; See Mfg Details	FLOOR	SEE MFG DETAILS	E 115V; DATA/COM
OPSVI	6008	1 3 O'Clock, Accessory Console 3 door unit Wall mounted;Closed Op; Owner Preferred Dispensers ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6008c	1 9 O'Clock, Accessory Console With Sink; Closed Op-PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6009	1 12 O'Clock Dental Cabinetry: ear Delivery Sys-Dental & Dental Assist Control,PC Mount & Access: ADEC	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVCI	6019	1 Chair, Dental 511; With Chair Mount and Monitor Mount: PER ADEC DESIGN	floor	PER ADEC DESIGN DRWGS	SEE ADEC DESIGN DRWGS
OPSVI	6060a	1 Scaler, Dental: Owner to Specify	Attached	13x15x4	SEE DESIGN DRWGS
OPSVI	6063	1 Stool, Footplate,Adj.Ht.,(Dental Assistant): ADEC	SEE ADEC DRWGS	22diam x36	PER ADEC DESIGN DRWGS
OPSVI	6064	1 Stool, Adj.Ht.,(Dentist): ADEC	floor	22diam x36	PER ADEC DESIGN DRWGS
OPSVCI	6065	1 Switch, Additional; Tech Use for Radiography Exposure: BENCO SEE MFG DETAILS	wall	3"w x 4"h	SEE BENCO/PLAN MECA DESIGN DRWGS
OPVI	8700	1 Monitor, Vital Signs, BP, Temp, Pulse, Optional SPO2, Portable On Stand	stand	11"wx10"hx6"d w/stand21" diameter x 44"h	E 115V
OPSVCI	9907	1 Nitrous Supply System;accepts an IEC/EN 60320-1 C13 connector; DETAILS BY BENCO	cabinetry mount	8"w x4"d x7.5"h	Nitrous/O2;IEC/EN 60320-1 C14

DEPARTMENT TITLE: CIRCULATION-2ND FLOOR

ROOM NO 02-18-00

RMTITLE CIRCULATION, 2ND FLOOR (AED LOCATION)

<u>GROUP</u>	<u>EQNO</u>	<u>QTY</u>	<u>DESCRIPTION</u>	<u>LOCATION</u>	<u>SIZE</u>	<u>UTILITIES</u>
OPOI	8202	1	Defibrillator, External, Automated (AED) *TO BE STORED INSIDE OF 8202a	stored	8.9"w x7.1"h x2.4"d	None
OPCI	8202a	1	Cabinet, Semi Recessed for AED Defibrillator: PROVIDED & INSTALLED BY CONTRACTOR	recessed/w all	see mfr details	battery operated